

SAMPLE 検査成績表
<TEST REPORT>

1/1

品名<Model Name> (クランプオンAC/DC センサ<CLAMP ON AC/DC SENSOR>)
 形名<Model Number> (9274)
 製造番号<Serial No.> (No. 110812345)
 検査年月日<Test Date> (2011 - 08 - 08)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (50Hz)

項目 <Item>	検査点 <Test Point>	出力基底値 <Fiducial Value>	許容範囲 <Tolerance>	出力値 <Output>
確 度 <Accuracy>	AC 19.00 A	AC 1.9000 V	1.8885 V ~ 1.9115 V	(1.9000 V)
	10.00 A	1.0000 V	0.9930 V ~ 1.0070 V	(1.0000 V)
	5.000 A	0.5000 V	0.4955 V ~ 0.5045 V	(0.5000 V)
	2.000 A	0.2000 V	0.1970 V ~ 0.2030 V	(0.2000 V)
	1.000 A	100.0 mV	97.5 mV ~ 102.5 mV	(100.0 mV)
	0.500 A	50.0 mV	47.8 mV ~ 52.2 mV	(50.0 mV)
	0.200 A	20.0 mV	17.9 mV ~ 22.1 mV	(20.0 mV)
	0.100 A	10.0 mV	8.0 mV ~ 12.0 mV	(10.0 mV)
	DC +19.00 A	DC +1.9000V	+1.8885V ~ +1.9115V	(+ 1.9000 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()