

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (ネットワークD <NETWORK D>)
形名<Model Number> (9499)
製造番号<Serial No.> (No. 030512945)
検査年月日<Test Date> (2011-07-26)
(<YYYY-MM-DD>)
試験条件<Test Conditions> (25.5 °C, 58 %rh)

項目 <Item>	入力 <Input>	周波数 <Freq.>	抵抗設定 <Resistance>	許容範囲 <Tolerance>			出力値 <Output Value>	
確度(フィルタ特性)								
<Accuracy>	1mA	DC	1.0k Ω	0.9950 V	~	1.0050 V	(	1.0000 V)
	1mA	DC	1.5k Ω	1.4925 V	~	1.5075 V	(	1.5000 V)
	1mA	DC	2.0k Ω	1.9900 V	~	2.0100 V	(	2.0000 V)
	インピーダンス <Impedance *1>	50Hz	1.5k Ω //0.15uF	1488.7 Ω	~	1503.7 Ω	(	1500.0 Ω)
	〃	100Hz	1.5k Ω //0.15uF	1477.8 Ω	~	1492.6 Ω	(	1490.0 Ω)
	〃	300Hz	1.5k Ω //0.15uF	1374.0 Ω	~	1387.8 Ω	(	1380.0 Ω)
	〃	1kHz	1.5k Ω //0.15uF	847.9 Ω	~	884.5 Ω	(	880.0 Ω)
	〃	3kHz	1.5k Ω //0.15uF	336.4 Ω	~	352.1 Ω	(	350.0 Ω)
	〃	10kHz	1.5k Ω //0.15uF	102.72 Ω	~	108.95 Ω	(	105.00 Ω)
	〃	30kHz	1.5k Ω //0.15uF	33.65 Ω	~	37.06 Ω	(	37.00 Ω)
	〃	100kHz	1.5k Ω //0.15uF	9.40 Ω	~	11.82 Ω	(	11.00 Ω)

(インピーダンスはヒューズ両端短絡時)
(*1 Fuse terminals shorted)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.9499-3