

SAMPLE 検査成績表
<TEST REPORT>

1/1

品名<Model Name> (外部信号入力ユニット<EXTERNAL SIGNAL INPUT UNIT>)
 形名<Model Number> (9603)
 製造番号<Serial No.> (No. 100524545)
 検査年月日<Test Date> (2010-05-26)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (24.6 °C, 56 %rh)

項目 <Item>	入力モード <Input Mode>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>					
ch A	アナログ	10V	+10V	+9.980 V ~ +10.020 V	(+9.999 V)
	<ANALOG>	5V	+5V	+4.9900 V ~ +5.0100 V	(+4.9999 V)
		1V	+1V	+0.9980 V ~ +1.0020 V	(+1.0000 V)
			+0.5V	+0.4985 V ~ +0.5015 V	(+0.5000 V)
			+0.05V	+0.0490 V ~ +0.0510 V	(+0.0499 V)
			-0.05V	-0.0510 V ~ -0.0490 V	(-0.0499 V)
			0V	-0.0010 V ~ +0.0010 V	(+0.0000 V)
ch B	アナログ	10V	+10V	+9.980 V ~ +10.020 V	(+10.000 V)
	<ANALOG>	5V	+5V	+4.9900 V ~ +5.0100 V	(+5.0000 V)
		1V	+1V	+0.9980 V ~ +1.0020 V	(+1.0000 V)
			+0.5V	+0.4985 V ~ +0.5015 V	(+0.4999 V)
			+0.05V	+0.0490 V ~ +0.0510 V	(+0.0497 V)
			-0.05V	-0.0510 V ~ -0.0490 V	(-0.0501 V)
			0V	-0.0010 V ~ +0.0010 V	(+0.0000 V)
周波数<Frequency>					
ch B	パルス<PULSE>	50Hz	0/5V 50Hz	49.949 Hz ~ 50.051 Hz	(50.001 Hz)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.9603-2