

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (外部信号入力ユニット<EXTERNAL SIGNAL INPUT UNIT>)
形名<Model Number> (9603-01)
製造番号<Serial No.> (No. 070805494)
検査年月日<Test Date> (2009-08-18)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.2 °C, 58 %rh)

項目 ＜Item＞	入力モード ＜Input mode＞	レンジ ＜Range＞	入力 ＜Input＞	許容範囲 ＜Tolerance＞		表示値 ＜Indicated Value＞	
確度＜Accuracy＞							
ch A	—	2mV/V	4000uε	+1.9967 N・m	～ +2.0033 N・m	(	+1.9994 N・m)
歪ゲージ		1.5mV/V	3000uε	+1.4976 N・m	～ +1.5024 N・m	(	+1.4988 N・m)
＜STRAIN GAGE＞		1mV/V	2000uε	+0.9984 N・m	～ +1.0016 N・m	(	+0.9999 N・m)
			100uε	+0.0493 N・m	～ +0.0507 N・m	(	+0.0501 N・m)
ch A	アナログ	10V	+10V	+9.980 V	～ +10.020 V	(	+9.999 V)
	＜ANALOG＞	5V	+5V	+4.9900 V	～ +5.0100 V	(	+5.0002 V)
		1V	+1V	+0.9980 V	～ +1.0020 V	(	+1.0000 V)
			+0.5V	+0.4985 V	～ +0.5015 V	(	+0.5000 V)
			+0.05V	+0.0490 V	～ +0.0510 V	(	+0.0501 V)
			-0.05V	-0.0510 V	～ -0.0490 V	(	-0.0499 V)
			0V	-0.0010 V	～ +0.0010 V	(	0.0000 V)
ch B	アナログ	10V	+10V	+9.980 V	～ +10.020 V	(	+9.999 V)
	＜ANALOG＞	5V	+5V	+4.9900 V	～ +5.0100 V	(	+4.9999 V)
		1V	+1V	+0.9980 V	～ +1.0020 V	(	+1.0001 V)
			+0.5V	+0.4985 V	～ +0.5015 V	(	+0.5001 V)
			+0.05V	+0.0490 V	～ +0.0510 V	(	+0.0501 V)
			-0.05V	-0.0510 V	～ -0.0490 V	(	-0.0497 V)
			0V	-0.0010 V	～ +0.0010 V	(	0.0000 V)
周波数＜Frequency＞							
ch B	パルス＜PULSE＞	50Hz	0/5V 50Hz	49.949 Hz	～ 50.051 Hz	(	50.001 Hz)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()