

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (クランプオンAC/DCセンサ<CLAMP ON AC/DC SENSOR>)
形名<Model Number> (9692)
製造番号<Serial No.> (No. 120712342)
検査年月日<Test Date> (2012 - 07 - 31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.2 °C, 50 %rh)

項目 <Item>	入力値 <Input Value>	出力基底値 <Fiducial Value>	許容範囲 <Tolerance>	出力値 <Output Value>
--------------	----------------------	---------------------------	---------------------	-----------------------

振幅確度
<Amplitude accuracy>

AC	180 A 55Hz	AC 1800 mV	1781.5 mV ~ 1818.5 mV	(1799.9 mV)
AC	18 A 55Hz	AC 180 mV	177.70 mV ~ 182.30 mV	(179.99 mV)
DC	180 A	DC 1800 mV	1781.5 mV ~ 1818.5 mV	(1804.5 mV)
DC	-180 A	DC -1800 mV	-1818.5 mV ~ -1781.5 mV	(-1804.4 mV)
DC	18 A	DC 180 mV	177.70 mV ~ 182.30 mV	(180.53 mV)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()