

SAMPLE
検査成績表
<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

品名<Model Name> (クランプオンノイズセンサ)
<CLAMP ON NOISE SENSOR>
形名<Model Number> (9754)
製造番号<Serial No.> (No. 190712345)
検査年月日<Test Date> (2019-07-19)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	入力 <Input>			出力基底値 <Fiducial Value>			許容範囲 <Tolerance>			校正値 * <Calibration Value>				
1. 振幅確度<Amplitude Accuracy>														
	AC	10 A	15kHz	AC	1	V	0.96999	V	~	1.03001	V (	1.00000	V)	
	AC	1 A	15kHz	AC	100	mV	96.99	mV	~	103.01	mV (	100.00	mV)	
	AC	0.1 A	15kHz	AC	10	mV	9.69	mV	~	10.31	mV (	10.00	mV)	
	AC	10 mA	15kHz	AC	1	mV	0.960	mV	~	1.040	mV (	1.000	mV)	
項目 <Item>							許容範囲 <Tolerance>			結果 * <Result>				
2.周波数帯域<BandWidth>														
							1kHz ~ 100MHz (-3dB)			(			PASS	)

備考<Note>

* FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()