

## UK DECLARATION OF CONFORMITY

<b>Manufacturer's Name and address</b> /	HIOKI E.E. CORPORATION 81 Koizumi, Ueda, Nagano 386-1192, Japan
<b>Product Name(s)</b>	AC LEAKAGE CLAMP METER
<b>Product model(s)</b>	CM4001
<b>Accessories and Option</b>	<b>All accessories and options of above product(s), if applicable.</b>

<b>UK legislation</b>		<b>Dedicated standard</b>
<b>Safety</b>	Electrical Equipment (Safety) Regulations 2016 No.1101 [as amended]	EN 61010-2-032:2012
<b>EMC</b>	Electromagnetic Compatibility Regulations 2016 No.1091 [as amended]	EN 61326-1:2013
<b>RoHS</b>	Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 No.3032 [as amended]	EN IEC 63000:2018

**The undersigned hereby declares that the product(s) is/are in conformity with the above standard(s) and with all essential requirements of the above UK legislation.**

30/06/2021  
(DD/MM/YYYY)

HIOKI E.E. CORPORATION



\_\_\_\_\_  
Katsuhiro Takeuchi  
**Director of Quality Assurance**

CM4001A995-00