



Certificate of Compliance

Certificate: 80008923

Master Contract: 300342

Project: 80128128

Date Issued: May 26, 2022

Issued To: Hioki E.E. Corporation
81 Koizumi
Ueda, Nagano, 386-1192
Japan

Attention: Inoue Tomoyuki

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Jerry Villamar

PRODUCTS

CLASS - C363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Electrical Measurement and Test Equipment

CLASS - C363186 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

- Battery HiTester, Models: BT3562, BT3562-01, BT3563, BT3563-01, Detachable cord connected, rated: 100-240 V, 50/60Hz, 30 VA.
- Battery HiTester, Models: BT3561A, BT3562A, BT3563A, Detachable cord connected, rated: 100-240 V, 50/60Hz, 35VA

Notes:

1. The above models are Equipment Class I, Pollution Degree 2, Measurement Category None (Measurement Inputs) and Overvoltage category II (Mains input).
2. Mode of operation: Continuous
3. Environmental Conditions: Normal: 0 °C to 40 °C, 80 % RH or less, non-condensing, 2000 m max.



Certificate: 80008923
Project: 80128128

Master Contract: 300342
Date Issued: 2022-05-26

APPLICABLE REQUIREMENTS

CSA Standards:

- | | |
|--|--|
| CAN/CSA C22.2 No. 61010-1-12, UPD1: 2015, UPD2: 2016, AMD1: 2018 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| CSA-C22.2 No. 61010-2-030-18 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits |

UL Standards:

- | | |
|--|--|
| UL 61010-1, 3 rd edition (2012), AMD1: 2018 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |
| UL Std. No. 61010-2-030 (2 nd Edition) | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits |

CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) Equipment is not to be used with flammable liquids.

Notes:

Products certified under Class C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80008923

Master Contract: 300342

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80128128	2022-05-26	Update report: 80073886 to update the PCB for BT3561A, BT3561B and BT3563A information to Interchangeable INT).
80073886	2021-05-20	Update report: 80008923 to include new models: BT-3561A, BT3562A & BT3563A and to add alternate construction.
80008923	2020-02-11	Battery HiTester, Model BT3562, BT3562-01, BT3563, BT3563-01 (CAN/US)