

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (パワーアナライザ<POWER ANALYZER>)
 形名<Model Number> (3390-10)
 製造番号<Serial No.> (No. 0)
 検査年月日<Test Date> (1900-01-00)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (0.0 °C, 0 %rh)

1. 電圧精度<Voltage Accuracy>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *1,*3			
			<Volt>	<Freq>		CH1	CH2	CH3	CH4
電圧 <Voltage>	AC+DC	1.5kV	1kV	55Hz	0.9988 kV ~ 1.0012 kV (1.0000 kV)	(1.0000 kV)	(1.0000 kV)	(1.0000 kV)	(1.0000 kV)
			0V	-	0.0000 kV ~ 0.0000 kV (0.0000 kV)	(0.0000 kV)	(0.0000 kV)	(0.0000 kV)	(0.0000 kV)
		600V	600V	55Hz	599.40 V ~ 600.60 V (600.00 V)	(600.00 V)	(600.00 V)	(600.00 V)	(600.00 V)
		300V	300V	55Hz	299.70 V ~ 300.30 V (300.00 V)	(300.00 V)	(300.00 V)	(300.00 V)	(300.00 V)
		150V	150V	55Hz	149.85 V ~ 150.15 V (150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)
			100V	55Hz	99.88 V ~ 100.12 V (100.00 V)	(100.00 V)	(100.00 V)	(100.00 V)	(100.00 V)
			75V	55Hz	74.89 V ~ 75.11 V (75.00 V)	(75.00 V)	(75.00 V)	(75.00 V)	(75.00 V)
			1.5V	55Hz	1.43 V ~ 1.57 V (1.50 V)	(1.50 V)	(1.50 V)	(1.50 V)	(1.50 V)
	0V		-	0.00 V ~ 0.00 V (0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	
	60V		60V	55Hz	59.940 V ~ 60.060 V (60.000 V)	(60.000 V)	(60.000 V)	(60.000 V)	(60.000 V)
	30V	30V	55Hz	29.970 V ~ 30.030 V (30.000 V)	(30.000 V)	(30.000 V)	(30.000 V)	(30.000 V)	
	15V	15V	55Hz	14.985 V ~ 15.015 V (15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)	
	0V	-	0.000 V ~ 0.000 V (0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)	
	AC	1.5kV	1kV	55Hz	0.9988 kV ~ 1.0012 kV (1.0000 kV)	(1.0000 kV)	(1.0000 kV)	(1.0000 kV)	(1.0000 kV)
			600V	600V	55Hz	599.40 V ~ 600.60 V (600.00 V)	(600.00 V)	(600.00 V)	(600.00 V)
		300V	300V	55Hz	299.70 V ~ 300.30 V (300.00 V)	(300.00 V)	(300.00 V)	(300.00 V)	(300.00 V)
150V		150V	55Hz	149.85 V ~ 150.15 V (150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	
		60V	60V	55Hz	59.940 V ~ 60.060 V (60.000 V)	(60.000 V)	(60.000 V)	(60.000 V)	(60.000 V)
		30V	30V	55Hz	29.970 V ~ 30.030 V (30.000 V)	(30.000 V)	(30.000 V)	(30.000 V)	(30.000 V)
		15V	15V	55Hz	14.985 V ~ 15.015 V (15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)
		DC	1.5kV	+1kV		0.9975 kV ~ 1.0025 kV (1.0000 kV)	(1.0000 kV)	(1.0000 kV)	(1.0000 kV)
	0V				0.0000 kV ~ 0.0000 kV (0.0000 kV)	(0.0000 kV)	(0.0000 kV)	(0.0000 kV)	(0.0000 kV)
-1kV			-1.0025 kV ~ -0.9975 kV (-1.0000 kV)	(-1.0000 kV)	(-1.0000 kV)	(-1.0000 kV)	(-1.0000 kV)		
600V	+600V			598.80 V ~ 601.20 V (600.00 V)	(600.00 V)	(600.00 V)	(600.00 V)	(600.00 V)	
	0V			0.00 V ~ 0.00 V (0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	
300V	+300V			299.40 V ~ 300.60 V (300.00 V)	(300.00 V)	(300.00 V)	(300.00 V)	(300.00 V)	
	0V			0.00 V ~ 0.00 V (0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	
150V	+150V			149.70 V ~ 150.30 V (150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	
	0V		0.00 V ~ 0.00 V (0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)	(0.00 V)		
-150V			-150.30 V ~ -149.70 V (-150.00 V)	(-150.00 V)	(-150.00 V)	(-150.00 V)	(-150.00 V)		
	60V	+60V		59.880 V ~ 60.120 V (60.000 V)	(60.000 V)	(60.000 V)	(60.000 V)	(60.000 V)	
0V			0.000 V ~ 0.000 V (0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)		
30V	+30V		29.940 V ~ 30.060 V (30.000 V)	(30.000 V)	(30.000 V)	(30.000 V)	(30.000 V)		
	0V		0.000 V ~ 0.000 V (0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)		
15V	+15V		14.970 V ~ 15.030 V (15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)		
	+0.15V		0.135 V ~ 0.165 V (0.150 V)	(0.150 V)	(0.150 V)	(0.150 V)	(0.150 V)		
-15V			-15.030 V ~ -14.970 V (-15.000 V)	(-15.000 V)	(-15.000 V)	(-15.000 V)	(-15.000 V)		
	0V		0.000 V ~ 0.000 V (0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)	(0.000 V)		

備考<Note>

*電流・電力検査における電流入力は、クランプオンセンサの入力に比例した電圧入力で行います。

<On current / power inspection, current input value is indicated and executed by voltage in proportion to the input for Clamp-on Sensor.>

*特に設定の明記のない場合、以下の設定で検査しています。

センサ:「50A定格」、LPF:「OFF」、ゼロクロスフィルタ:「弱」、アベレージ:「OFF」

<Inspection settings are as follows except noted.>

CT sensor: 50A rated, LPF: OFF, Zero-cross filter: Low, Averaging: OFF>

*1. ゼロサプレス設定 : 0.5%fs. <Zero-suppress setting: 0.5% f.s.>

*2. ゼロサプレス設定 : OFF <Zero-suppress setting: OFF>

*3. FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

検査成績表
(TEST REPORT)

製造番号<Serial No.>
検査年月日<Test Date>

(No. 0)
(1900-01-00)
<YYYY-MM-DD>

2. 電流精度<Current Accuracy>

項目 <Item>	レクティファイ <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *1,*3			
			<Curr>	<Freq>		CH1	CH2	CH3	CH4
電流 <Current>	AC+DC	50A	50A	55Hz	49.950 A ~ 50.050 A	(50.000 A)	(50.000 A)	(50.000 A)	(50.000 A)
			0A	-	0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	(0.000 A)	(0.000 A)
		20A	20A	55Hz	19.980 A ~ 20.020 A	(20.000 A)	(20.000 A)	(20.000 A)	(20.000 A)
			10A	55Hz	9.990 A ~ 10.010 A	(10.000 A)	(10.000 A)	(10.000 A)	(10.000 A)
			5A	55Hz	4.9950 A ~ 5.0050 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)
		2.5A	55Hz	2.4963 A ~ 2.5037 A	(2.5000 A)	(2.5000 A)	(2.5000 A)	(2.5000 A)	
			0.5A	55Hz	0.4973 A ~ 0.5027 A	(0.5000 A)	(0.5000 A)	(0.5000 A)	(0.5000 A)
			0.05A	55Hz	0.0475 A ~ 0.0525 A	(0.0500 A)	(0.0500 A)	(0.0500 A)	(0.0500 A)
			0A	-	0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	(0.0000 A)	(0.0000 A)
			2A	55Hz	1.9980 A ~ 2.0020 A	(2.0000 A)	(2.0000 A)	(2.0000 A)	(2.0000 A)
	1A	1A	55Hz	0.9990 A ~ 1.0010 A	(1.0000 A)	(1.0000 A)	(1.0000 A)	(1.0000 A)	
		0A	-	0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	(0.0000 A)	(0.0000 A)	
	AC	50A	50A	55Hz	49.950 A ~ 50.050 A	(50.000 A)	(50.000 A)	(50.000 A)	(50.000 A)
			20A	55Hz	19.980 A ~ 20.020 A	(20.000 A)	(20.000 A)	(20.000 A)	(20.000 A)
		10A	55Hz	9.990 A ~ 10.010 A	(10.000 A)	(10.000 A)	(10.000 A)	(10.000 A)	
		5A	55Hz	4.9950 A ~ 5.0050 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)	
		2A	55Hz	1.9980 A ~ 2.0020 A	(2.0000 A)	(2.0000 A)	(2.0000 A)	(2.0000 A)	
		1A	55Hz	0.9990 A ~ 1.0010 A	(1.0000 A)	(1.0000 A)	(1.0000 A)	(1.0000 A)	
	DC	50A	+50A	49.900 A ~ 50.100 A	(50.000 A)	(50.000 A)	(50.000 A)	(50.000 A)	
			0A	0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	(0.000 A)	(0.000 A)	
-50A			-50.100 A ~ -49.900 A	(-50.000 A)	(-50.000 A)	(-50.000 A)	(-50.000 A)		
20A			+20A	19.960 A ~ 20.040 A	(20.000 A)	(20.000 A)	(20.000 A)	(20.000 A)	
10A		0A	0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	(0.000 A)	(0.000 A)		
		+10A	9.980 A ~ 10.020 A	(10.000 A)	(10.000 A)	(10.000 A)	(10.000 A)		
5A		0A	0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	(0.000 A)	(0.000 A)		
		+5A	4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)		
2A		0A	0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	(0.0000 A)	(0.0000 A)		
		+2A	1.9960 A ~ 2.0040 A	(2.0000 A)	(2.0000 A)	(2.0000 A)	(2.0000 A)		
1A		0A	0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	(0.0000 A)	(0.0000 A)		
		+1A	0.9980 A ~ 1.0020 A	(1.0000 A)	(1.0000 A)	(1.0000 A)	(1.0000 A)		
		+0.01A	0.0090 A ~ 0.0110 A	(0.0100 A)	(0.0100 A)	(0.0100 A)	(0.0100 A)		
		0A	0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	(0.0000 A)	(0.0000 A)		
		-1A	-1.0020 A ~ -0.9980 A	(-1.0000 A)	(-1.0000 A)	(-1.0000 A)	(-1.0000 A)		

3. 電力精度<Power Accuracy>

項目 <Item>	レクティファイ <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *1,*3			
			<Volt/Curr/Pow/Freq/PF>			CH1	CH2	CH3	CH4
-1. 力率 1 (cos φ=1)	AC+DC	150V 5A	150V, 5A, 750W, 55Hz	1	749.25 W ~ 750.75 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)
			150V, 2.5A, 375W, 55Hz	1	374.44 W ~ 375.56 W	(375.00 W)	(375.00 W)	(375.00 W)	(375.00 W)
		150V, 0.5A, 75W, 55Hz	1	74.59 W ~ 75.41 W	(75.00 W)	(75.00 W)	(75.00 W)	(75.00 W)	
		150V, 0.05A, 7.5W, 55Hz	1	7.13 W ~ 7.87 W	(7.50 W)	(7.50 W)	(7.50 W)	(7.50 W)	
		150V(55Hz), 0A, 0W	-	0.00 W ~ 0.00 W	(0.00 W)	(0.00 W)	(0.00 W)	(0.00 W)	
	1500V 50A	1000V, 50A, 50kW, 55Hz	1	49.938 kW ~ 50.062 kW	(50.000 kW)	(50.000 kW)	(50.000 kW)	(50.000 kW)	
	1500V 1A	1000V, 1A, 1kW, 55Hz	1	0.9988 kW ~ 1.0012 kW	(1.0000 kW)	(1.0000 kW)	(1.0000 kW)	(1.0000 kW)	
	15V 50A	15V, 50A, 750W, 55Hz	1	749.25 W ~ 750.75 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)	
	15V 1A	15V, 1A, 15W, 55Hz	1	14.985 W ~ 15.015 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)	
			15V(55Hz), 0A, 0W	-	0.000 W ~ 0.000 W	(0.000 W)	(0.000 W)	(0.000 W)	(0.000 W)
DC	15V 1A	+15V, +1A, +15W	14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)		
		-15V, -1A, +15W	14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)		
		+15V, +0.01A, +0.15W	0.135 W ~ 0.165 W	(0.150 W)	(0.150 W)	(0.150 W)	(0.150 W)		
	0V, 0A, 0W	0.000 W ~ 0.000 W	(0.000 W)	(0.000 W)	(0.000 W)	(0.000 W)			
	1500V 50A	+1000V, +50A, +50kW	49.875 kW ~ 50.125 kW	(50.000 kW)	(50.000 kW)	(50.000 kW)	(50.000 kW)		
		-1000V, -50A, +50kW	49.875 kW ~ 50.125 kW	(50.000 kW)	(50.000 kW)	(50.000 kW)	(50.000 kW)		

SAMPLE

検査成績表

<TEST REPORT>

製造番号<Serial No.> (No. 0)
 検査年月日<Test Date> (1900-01-00)
 <YYYY-MM-DD>

3. 電力精度<Power Accuracy>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *2,*3			
			<Volt/Curr/Pow/Freq/PF>			CH1	CH2	CH3	CH4
-2. 力率 0 <cos φ=0>	AC+DC	150V 5A	150V, 5A, 0W, 55Hz, LEAD 0		-1.12 W ~ 1.12 W	(0.00 W)	(0.00 W)	(0.00 W)	(0.00 W)
			150V, 5A, 0W, 55Hz, LAG 0		-1.12 W ~ 1.12 W	(0.00 W)	(0.00 W)	(0.00 W)	(0.00 W)
			15V 50A 15V, 50A, 0W, 55Hz, LEAD 0		-1.12 W ~ 1.12 W	(0.00 W)	(0.00 W)	(0.00 W)	(0.00 W)
			15V, 50A, 0W, 55Hz, LAG 0		-1.12 W ~ 1.12 W	(0.00 W)	(0.00 W)	(0.00 W)	(0.00 W)

4. ピーク測定<Peak Measurement>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *3			
			<Volt/Curr>	<Freq>		CH1	CH2	CH3	CH4
-1. 電圧 <Voltage>	AC+DC	15V	+15V	DC	14.700 V ~ 15.300 V	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)
			-15V	DC	-15.300 V ~ -14.700 V	(-15.000 V)	(-15.000 V)	(-15.000 V)	(-15.000 V)
-2. 電流 <Current>	AC+DC	1A	+1A	DC	0.9800 A ~ 1.0200 A	(1.0000 A)	(1.0000 A)	(1.0000 A)	(1.0000 A)
			-1A	DC	-1.0200 A ~ -0.9800 A	(-1.0000 A)	(-1.0000 A)	(-1.0000 A)	(-1.0000 A)

5. 周波数特性<Frequency Characteristic>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *3							
			<Volt/Curr>	<Freq>		CH1	CH2	CH3	CH4				
-1. 電圧 <Voltage>	AC+DC	150V	150V	10Hz	149.55 V ~ 150.45 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				50Hz	149.85 V ~ 150.15 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				60Hz	149.85 V ~ 150.15 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				100Hz	149.70 V ~ 150.30 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				500Hz	149.70 V ~ 150.30 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				1kHz	149.70 V ~ 150.30 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				5kHz	149.55 V ~ 150.45 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				10kHz	149.55 V ~ 150.45 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				100kHz	148.05 V ~ 151.95 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)				
				MEAN	150V	150V	60Hz	149.85 V ~ 150.15 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	
							1kHz	149.70 V ~ 150.30 V	(150.00 V)	(150.00 V)	(150.00 V)	(150.00 V)	
				-2. 電流 <Current>	AC+DC	5A	5A	10Hz	4.9850 A ~ 5.0150 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)
								50Hz	4.9950 A ~ 5.0050 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)
								60Hz	4.9950 A ~ 5.0050 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)
100Hz	4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)					(5.0000 A)	(5.0000 A)				
500Hz	4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)					(5.0000 A)	(5.0000 A)				
1kHz	4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)					(5.0000 A)	(5.0000 A)				
5kHz	4.9850 A ~ 5.0150 A	(5.0000 A)	(5.0000 A)					(5.0000 A)	(5.0000 A)				
10kHz	4.9850 A ~ 5.0150 A	(5.0000 A)	(5.0000 A)					(5.0000 A)	(5.0000 A)				
100kHz	4.9350 A ~ 5.0650 A	(5.0000 A)	(5.0000 A)					(5.0000 A)	(5.0000 A)				
MEAN	5A	5A	60Hz					4.9950 A ~ 5.0050 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)	
			1kHz					4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)	
-3. 電力 <Power>	AC+DC	150V 5A	150V, 5A, 750W					10Hz	747.75 W ~ 752.25 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)
								50Hz	749.25 W ~ 750.75 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)
								60Hz	749.25 W ~ 750.75 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)
				100Hz	748.50 W ~ 751.50 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)				
				500Hz	748.50 W ~ 751.50 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)				
				1kHz	748.50 W ~ 751.50 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)				
				5kHz	747.75 W ~ 752.25 W	(750.00 W)	(750.00 W)	(750.00 W)	(750.00 W)				
				10Hz	14.955 W ~ 15.045 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)				
				50Hz	14.985 W ~ 15.015 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)				
	60Hz	14.985 W ~ 15.015 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)							
	100Hz	14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)							
	500Hz	14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)							
	1kHz	14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)							
	5kHz	14.955 W ~ 15.045 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)							
	AC+DC	15V 1A	15V, 1A, 15W	10Hz	14.955 W ~ 15.045 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)				
				50Hz	14.985 W ~ 15.015 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)				
				60Hz	14.985 W ~ 15.015 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)				
				100Hz	14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)				
500Hz				14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)					
1kHz				14.970 W ~ 15.030 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)					
5kHz				14.955 W ~ 15.045 W	(15.000 W)	(15.000 W)	(15.000 W)	(15.000 W)					

検査成績表
TEST REPORT

製造番号<Serial No.> (No. 0)
 検査年月日<Test Date> (1900-01-00)
 <YYYY-MM-DD>

6. LPF精度<LPF Accuracy>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力		許容範囲 <Tolerance>	測定値<Measured Value> *3			
			<Volt/Curr>	<Freq>		CH1	CH2	CH3	CH4
-1.電圧 <Voltage>	AC+DC	15V	500Hz	15V 60Hz	14.970 V ~ 15.030 V	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)
			5kHz	15V 500Hz	14.970 V ~ 15.030 V	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)
			100kHz	15V 20kHz	14.925 V ~ 15.075 V	(15.000 V)	(15.000 V)	(15.000 V)	(15.000 V)
-2.電流 <Current>	AC+DC	5A	500Hz	5A 60Hz	4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)
			5kHz	5A 500Hz	4.9900 A ~ 5.0100 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)
			100kHz	5A 20kHz	4.9750 A ~ 5.0250 A	(5.0000 A)	(5.0000 A)	(5.0000 A)	(5.0000 A)

7. 周波数測定<Frequency Measurement>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	測定値<Measured Value> *3			
			<Volt/Curr>	<Freq>		CH1	CH2	CH3	CH4
	AC+DC	15V	15V	55Hz	54.972 Hz ~ 55.028 Hz	(55.000 Hz)	(55.000 Hz)	(55.000 Hz)	(55.000 Hz)
		1A	1A	55Hz	54.972 Hz ~ 55.028 Hz	(55.000 Hz)	(55.000 Hz)	(55.000 Hz)	(55.000 Hz)