

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E E CORPORATION

1/2

品名<Model Name> (マイクロハイコーダ<MICRO HiCORDER>)
 形名<Model Number> (8205-10)
 製造番号<Serial No.> (No.110812345)
 検査年月日<Test Date> (2011-08-18)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 電源周波数<Supply Frequency> (60Hz)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	標準器指示値 <Standard>
確度<Accuracy>				
DC V				
	0.1 V	0.1 V	0.0980 V ~ 0.1020 V	(0.1000 V)
	0.2 V	0.2 V	0.1960 V ~ 0.2040 V	(0.2000 V)
	0.5 V	0.5 V	0.490 V ~ 0.510 V	(0.500 V)
	1 V	1 V	0.980 V ~ 1.020 V	(1.000 V)
	2 V	2 V	1.960 V ~ 2.040 V	(2.000 V)
	5 V	5 V	4.90 V ~ 5.10 V	(5.00 V)
	10 V	10 V	9.80 V ~ 10.20 V	(10.00 V)
	20 V	20 V	19.60 V ~ 20.40 V	(20.00 V)
	50 V	50 V	49.0 V ~ 51.0 V	(50.0 V)
	100 V	100 V	98.0 V ~ 102.0 V	(100.0 V)
	200 V	200 V	196.0 V ~ 204.0 V	(200.0 V)
	500 V	500 V	490 V ~ 510 V	(500 V)
	10 V	-8 V	-8.20 V ~ -7.80 V	(-8.00 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

SAMPLE
検査成績表
 <TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

2/2

製造番号<Serial No.> (No.110812345)
 検査年月日<Test Date> (2011-08-18)
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	標準器指示値 <Standard>
--------------	----------------	---------------------	---------------------	----------------------

確度<Accuracy>

AC V

0.1 V	0.1 V/50Hz	0.0980 V	~ 0.1020 V	(0.1000 V)
0.2 V	0.2 V/50Hz	0.1960 V	~ 0.2040 V	(0.2000 V)
0.5 V	0.5 V/50Hz	0.490 V	~ 0.510 V	(0.500 V)
1 V	1 V/50Hz	0.980 V	~ 1.020 V	(1.000 V)
2 V	2 V/50Hz	1.960 V	~ 2.040 V	(2.000 V)
5 V	5 V/50Hz	4.90 V	~ 5.10 V	(5.00 V)
10 V	10 V/50Hz	9.80 V	~ 10.20 V	(10.00 V)
20 V	20 V/50Hz	19.60 V	~ 20.40 V	(20.00 V)
50 V	50 V/50Hz	49.0 V	~ 51.0 V	(50.0 V)
100 V	100 V/50Hz	98.0 V	~ 102.0 V	(100.0 V)
200 V	200 V/50Hz	196.0 V	~ 204.0 V	(200.0 V)
500 V	500 V/50Hz	490 V	~ 510 V	(500 V)

AC A

10 A	10 mA/50Hz	9.80 mA	~ 10.20 mA	(10.00 mA)
20 A	20 mA/50Hz	19.60 mA	~ 20.40 mA	(20.00 mA)
50 A	50 mA/50Hz	49.0 mA	~ 51.0 mA	(50.0 mA)
100 A	100 mA/50Hz	98.0 mA	~ 102.0 mA	(100.0 mA)
200 A	200 mA/50Hz	196.0 mA	~ 204.0 mA	(200.0 mA)
500 A	500 mA/50Hz	490 mA	~ 510 mA	(500 mA)
1000 A	500 mA/50Hz	470 mA	~ 530 mA	(500 mA)

備考<Note>

AC Aの値は、各レンジの定格値の1/1000を示す。

<The value of AC A indicates 1/1000 of rated value for each ranges.>