

**検査成績表**  
(TEST REPORT)

品名<Model Name> ( バッテリハイテスタ <BATTERY HiTESTER> )  
 形名<Model Number> ( BT3563 )  
 製造番号<Serial No.> ( No. 220112345 )  
 検査年月日<Test Date> ( 2022-01-01 )  
 ( <YYYY-MM-DD> )  
 検査条件<Test Conditions> ( 25.0 °C, 50 %rh )

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	(標準器校正値) <Std. Cal.Value>	許容範囲 <Tolerance>	校正値 *4 <Calibration Value>
1. 抵抗測定<Resistance Measurement> *1						
	Ω V	3m Ω	3 m Ω	( 3.0022 m Ω )	2.9862 m Ω ~ 3.0182 m Ω	( 3.0031 m Ω )
		30m Ω	30 m Ω	( 29.999 m Ω )	29.844 m Ω ~ 30.154 m Ω	( 29.988 m Ω )
		300m Ω	300 m Ω	( 299.99 m Ω )	298.44 m Ω ~ 301.54 m Ω	( 300.00 m Ω )
		3 Ω	3 Ω	( 2.9999 Ω )	2.9844 Ω ~ 3.0154 Ω	( 3.0000 Ω )
		30 Ω	30 Ω	( 30.003 Ω )	29.848 Ω ~ 30.158 Ω	( 30.008 Ω )
		300 Ω	300 Ω	( 299.99 Ω )	298.44 Ω ~ 301.54 Ω	( 300.01 Ω )
		3000 Ω	300 Ω	( 300.0 Ω )	298.0 Ω ~ 302.0 Ω	( 300.2 Ω )
		3000 Ω	3000 Ω	( 2997.7 Ω )	2982.2 Ω ~ 3013.2 Ω	( 2998.1 Ω )
2. 電圧測定<Voltage Measurement> *1						
	Ω V	6V	5.9 V	5.89938 V	~ 5.90062 V	( 5.89995 V )
			-5.9 V	-5.90062 V	~ -5.89938 V	( -5.89993 V )
		60V	59 V	58.9938 V	~ 59.0062 V	( 58.9986 V )
			-59 V	-59.0062 V	~ -58.9938 V	( -58.9986 V )
	*2	300V	290 V	289.968 V	~ 290.032 V	( 289.992 V )
			-290 V	-290.032 V	~ -289.968 V	( -289.993 V )
3. アナログ出力<Analog Output> *3						
			0 V	-0.0062 V	~ 0.0062 V	( - V )
			3 V	2.9938 V	~ 3.0062 V	( - V )

備考<Note>

\*1. サンプルングレート: SLOW  
<Sampling rate: SLOW>

\*2. BT3563, BT3563-01のレンジ構成となります。  
<The 300 V range is only available on the BT3563 or BT3563-01. >

\*3. BT3562-01, BT3563-01の機能となります。  
<The analog output function is only available on the BT3562-01 or BT3563-01>

\*4. FAIL判定箇所は、グレー表示としています。  
<FAIL judgment points are grayed out.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
( PASS )	( )	( )