

SAMPLE

## 検査成績表

&lt;TEST REPORT&gt;

HIOKI  
HIOKI E. E. CORPORATION

1/1

品名<Model Name> ( デジタルハイテスタ<DIGITAL HiTESTER> )  
 形名<Model Number> ( 3243 )  
 製造番号<Serial No.> ( No. 110712345 )  
 検査年月日<Test Date> ( 2011-12-31 )  
 ( <YYYY-MM-DD> )  
 検査条件<Test Conditions> ( 23.0 °C, 50 %rh )  
 検査周波数<Test Frequency> ( 50Hz )

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DCV	320 mV	300.0 mV	293.6 mV ~ 306.4 mV	( 293.6 mV )
	3.2 V	3.000 V	2.975 V ~ 3.025 V	( 2.975 V )
	32 V	30.00 V	29.57 V ~ 30.43 V	( 29.57 V )
	320 V	300.0 V	295.7 V ~ 304.3 V	( 295.7 V )
	450 V	450 V	441 V ~ 459 V	( 441 V )
ACV	3.2 V	3.000 V	2.923 V ~ 3.077 V	( 2.923 V )
	32 V	30.00 V	29.23 V ~ 30.77 V	( 29.23 V )
	320 V	300.0 V	292.3 V ~ 307.7 V	( 292.3 V )
	450 V	450 V	432 V ~ 468 V	( 432 V )
Ω	320 Ω	280.0 Ω	272.4 Ω ~ 287.6 Ω	( 272.4 Ω )
	3.2 kΩ	2.800 kΩ	2.738 kΩ ~ 2.862 kΩ	( 2.738 kΩ )
	32 kΩ	28.00 kΩ	27.38 kΩ ~ 28.62 kΩ	( 27.38 kΩ )
	320 kΩ	280.0 kΩ	273.8 kΩ ~ 286.2 kΩ	( 273.8 kΩ )
	3200 kΩ	2800 kΩ	2724 kΩ ~ 2876 kΩ	( 2724 kΩ )
	32 MΩ	28.00 MΩ	25.14 MΩ ~ 30.86 MΩ	( 25.14 MΩ )
Hz	3.2 kHz	3.000 kHz	2.996 kHz ~ 3.004 kHz	( 2.996 kHz )
	32 kHz	30.00 kHz	29.96 kHz ~ 30.04 kHz	( 29.96 kHz )
	320 kHz	300.0 kHz	299.6 kHz ~ 300.4 kHz	( 299.6 kHz )

## 備考&lt;Note&gt;

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
( PASS )	( )	( )

No.3243-5