

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (温度ハイテスタ <TEMPERATURE HiTESTER>)
 形名<Model Number> (3442)
 製造番号<Serial No.> (No. 2002-0237166)
 検査年月日<Test Date> (2006-03-13)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (22.5 °C, 50 %rh)

項目 <Item>	測定範囲 <Measurement Range>	試験点 <Test Point>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
センサ入力 <Sensor Input>				
	-100.0~	- 90 °C	-90.8 °C ~ -89.2 °C	(-90.4 °C)
	199.9°C	- 50 °C	-50.8 °C ~ -49.2 °C	(-50.3 °C)
		0 °C	-0.8 °C ~ 0.8 °C	(-0.4 °C)
		50 °C	49.2 °C ~ 50.8 °C	(49.7 °C)
		100 °C	99.1 °C ~ 100.9 °C	(99.7 °C)
		190 °C	189.1 °C ~ 190.9 °C	(189.8 °C)
	200~	200 °C	199 °C ~ 201 °C	(200 °C)
	1300°C	300 °C	299 °C ~ 301 °C	(300 °C)
		400 °C	399 °C ~ 401 °C	(400 °C)
		500 °C	498 °C ~ 502 °C	(500 °C)
		600 °C	598 °C ~ 602 °C	(600 °C)
		700 °C	698 °C ~ 702 °C	(700 °C)
		800 °C	798 °C ~ 802 °C	(800 °C)
		900 °C	898 °C ~ 902 °C	(900 °C)
		1000 °C	997 °C ~ 1003 °C	(1000 °C)
		1100 °C	1097 °C ~ 1103 °C	(1100 °C)
		1200 °C	1197 °C ~ 1203 °C	(1200 °C)

備考<Note>

温度はセンサの代わりに電圧発生器を接続し検査しています

<Temperature inspection is performed using the voltage generator instead of the temperature sensor.>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3442-4