

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (メグオームハイテスタ<MΩ HiTESTER>)
 形名<Model Number> (3450-12)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-03-12)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項 目 <Item>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	標準器指示値 <Standard>
確度 <Accuracy>			
MΩ	0.05 MΩ	0.0475 MΩ ~ 0.0525 MΩ	(0.0500 MΩ)
	1 MΩ	0.950 MΩ ~ 1.050 MΩ	(1.000 MΩ)
	20 MΩ	19.00 MΩ ~ 21.00 MΩ	(19.00 MΩ)
	30 MΩ	27.00 MΩ ~ 33.00 MΩ	(27.00 MΩ)
	50 MΩ	45.00 MΩ ~ 55.00 MΩ	(45.00 MΩ)
ACV			
(55Hz)	300 V	285.0 V ~ 315.0 V	(285.0 V)

項 目 <Item>	負荷抵抗 <Load Resistance>	許容範囲 <Tolerance>	標準器指示値 <Standard>
定格測定電流 <Rated Measurement Current>			
	0.25 MΩ	1.000 mA ~ 1.200 mA	(1.200 mA)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()