

SAMPLE 検査成績表
<TEST REPORT>

1/1

品名<Model Name> (液体試料用電極)
<LIQUID SAMPLE ELECTRODE>
形名<Model Number> (SME-8330)
製造番号<Serial No.> (123456789)
検査年月日<Test Date> (2011-12-25)
<YYYY-MM-DD>
検査条件<Test Conditions> (25.0 °C, 10 %rh)

項目 <Item>	許容範囲 <Tolerance>	* 測定値 <Measured Value>
1.絶縁抵抗 <Insulation Resistance>	印加電圧:DC1000V <Applied Voltage:DC1000V>	
-1.主電極 - 保護電極間 <Between Main electrode - Guard electrode>	1000000M Ω 以上 <1000000M Ω or more>	(200000 M Ω)
-2.保護電極間 - 対電極間 <Between Guard electrode - Counter electrode>	1000000M Ω 以上 <1000000M Ω or more>	(100000 M Ω)
2.主電極 - 対電極間容量 <Capacitance between main and counter electrode>	36.0 pF ~ 54.0 pF	(50 pF)
3.電極定数 <Electrode constant>	400 cm ~ 600 cm	(500 cm)

備考<Note>

* FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray>

総合判定<Overall Result> (PASS)	検査者<Inspected By> ()	承認者<Approved By> ()
----------------------------------	--------------------------	-------------------------