

SAMPLE

SR-2 STANDARD RESISTOR

INSPECTION DATA SHEET

Production No. 123456789

Inspected by _____

Approved by _____

Inspection Date May 13, 2016 Temperature 23.0°C Humidity 50%rh

1. Input voltage DC 10V

2. Mechanism inspection
Outward and Mechanism

Judgment	PASS
----------	------

3. Calibration Resistance degree of accuracy : $\pm 2\%$

Terminal	Test point	Nominal value	Calibration value	Synthetic value
0 ~ 10 MΩ	0 ~ 1MΩ	1 MΩ	1.000 MΩ	1.000 MΩ
	1 ~ 2.5 "	1.5 "	1.500 "	2.500 "
	2.5 ~ 5 "	2.5 "	2.500 "	5.000 "
	5 ~ 10 "	5 "	5.000 "	10.000 "
0 ~ 500MΩ	0 ~ 50MΩ	50 MΩ	50.00 MΩ	50.00 MΩ
	50 ~ 100 "	"	50.00 "	100.00 "
	100 ~ 150 "	"	50.00 "	150.0 "
	150 ~ 200 "	"	50.00 "	200.0 "
	200 ~ 250 "	"	50.00 "	250.0 "
	250 ~ 300 "	"	50.00 "	300.0 "
	300 ~ 350 "	"	50.00 "	350.0 "
	350 ~ 400 "	"	50.00 "	400.0 "
	400 ~ 450 "	"	50.00 "	450.0 "
450 ~ 500 "	"	50.00 "	500.0 "	
0~10000MΩ	0 ~ 1000MΩ	1000 MΩ	1000 MΩ	1000 MΩ
	1000 ~ 2000 "	"	1000 "	2000 "
	2000 ~ 3000 "	"	1000 "	3000 "
	3000 ~ 4000 "	"	1000 "	4000 "
	4000 ~ 5000 "	"	1000 "	5000 "
	5000 ~ 6000 "	"	1000 "	6000 "
	6000 ~ 7000 "	"	1000 "	7000 "
	7000 ~ 8000 "	"	1000 "	8000 "
	8000 ~ 9000 "	"	1000 "	9000 "
9000 ~ 10000 "	"	1000 "	10000 "	

Judgment	PASS
----------	------