

SAMPLE

検査成績表
<TEST REPORT>

品名<Model Name> (差動プローブ<DIFFERENTIAL PROBE>)
 形名<Model Number> (9322)
 製造番号<Serial No.> (No. 110509908)
 検査年月日<Test Date> (2011-05-17)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (24.7 °C, 52 %rh)

項目 <Item>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
DC振幅確度 <DC Amplitude Accuracy>			
	DC 1000V	DC 980mV ~ 1020mV	(1000 mV)
	DC -1000V	DC -1020mV ~ -980mV	(-1000 mV)
RMS確度 <RMS Accuracy>			
	AC 1000V/50Hz	DC 990mV ~ 1010mV	(1000 mV)
	AC 200V/100kHz	DC 160mV ~ 240mV	(200 mV)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------------	--------------------------	-------------------------