

SAMPLE

検査成績表
<TEST REPORT>

品名<Model Name> (ワイヤレスクランプロガー<WIRELESS CLAMP LOGGER>)
 形名<Model Number> (LR8513)
 製造番号<Serial No.> (No. 140747123)
 検査年月日<Test Date> (2020-02-03)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.2 °C, 52 %rh)

項目 <Item>	結果 <Result>
1. Bluetooth接続確認<Bluetooth Communication Check>	(PASS)
2. セルフチェック<Self Check>	
-1 ROMチェック<ROM Check>	(PASS)
-2. フラッシュROMチェック<FlashROM Check>	(PASS)
-3. FRAMチェック<FRAM Check>	(PASS)
-4. LCD チェック<LCD Check>	(PASS)
-5. キーチェック<Key Check>	(PASS)
3. 機能<Function>	
-1. 時計設定 <Clock Setting Check>	(PASS)
-2. 電池電圧検出 <Battery Voltage Detect Check>	(PASS)
-3. 電源表示チェック <Power Indication Check>	(PASS)
-4. Bluetooth電源チェック <Bluetooth Power Check>	(PASS)

項目 <Item>	設定<Setting> <Sensor(Range)/Filter>	入力CH <Input CH>	入力 *1 <Input>	許容範囲 <Tolerance>	校正値<Calibration Value> *2 CH1 CH2
4. 電流測定精度 <Current Measurement Accuracy>					
9675(500mA)	OFF	CH1	40.0mA, 60Hz	39.3mA ~ 40.7mA (40.2 mA)	
		CH2		39.3mA ~ 40.7mA	(40.2 mA)
		CH1	400.0mA, 50Hz	397.5mA ~ 402.5mA (400.1 mA)	
		CH2		397.5mA ~ 402.5mA	(400.1 mA)
		CH1	400.0mA, 60Hz	397.5mA ~ 402.5mA (400.2 mA)	
		CH2		397.5mA ~ 402.5mA	(400.1 mA)
	ON	CH1	400.0mA, 50Hz	397.5mA ~ 402.5mA (400.1 mA)	
			400.0mA, 60Hz	397.5mA ~ 402.5mA (400.2 mA)	
9675(5A)	OFF	CH1	0.040A, 60Hz	0.035A ~ 0.045A (0.043 A)	
			4.000A, 60Hz	3.975A ~ 4.025A (4.001 A)	
CT9692-90(20A)	ON	CH1	-16.00A, DC	-16.13A ~ -15.87A (-15.98 A)	
			-0.40A, DC	-0.45A ~ -0.35A (-0.39 A)	
			0.40A, DC	0.35A ~ 0.45A (0.42 A)	
			16.00A, DC	15.87A ~ 16.13A (16.00 A)	

備考<Note>

*1. 入力値は電流センサの入力に比例した電圧入力にて行います

<Input value is indicated and executed by voltage in proportion to the input for Current Sensor.>

*2. FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()