

SAMPLE 検査成績表
<TEST REPORT>

品名<Model Name> (AC/DC クランプメータ<AC/DC CLAMP METER>)
 形名<Model Number> (CM4371-50)
 製造番号<Serial No.> (No.211142205)
 検査年月日<Test Date> (2021-12-04)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (25.1°C, 21% RH)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
交流電流<AC Current>				
20.00A Filter OFF		1.00A 50Hz	0.91 A ~ 1.09 A	(0.99 A)
		18.00A 50Hz	17.69 A ~ 18.31 A	(17.98 A)
600.0A Filter OFF		10.0A 50Hz	9.6 A ~ 10.4 A	(10.1 A)
		540.0A 50Hz	532.7 A ~ 547.3 A	(539.7 A)
直流電流／交流＋直流電流<DC Current / AC+DC Current>				
20.00A Filter OFF		-18.00A DC	-18.31 A ~ -17.69 A	(-17.96 A)
		1.00A DC	0.91 A ~ 1.09 A	(0.98 A)
		18.00A DC	17.69 A ~ 18.31 A	(18.02 A)
600.0A Filter OFF		-540.0A DC	-547.3 A ~ -532.7 A	(-539.8 A)
		10.0A DC	9.6 A ~ 10.4 A	(10.0 A)
		540.0A DC	532.7 A ~ 547.3 A	(539.6 A)
交流電流Peak<AC Current Peak>				
20.00A		±142.5Apeak 50Hz	140.0 A ~ 145.0 A	(142.6 A)
			-145.0 A ~ -140.0 A	(-142.7 A)
600.0A		±1425Apeak 50Hz	1347 A ~ 1503 A	(1433 A)
			-1503 A ~ -1347 A	(-1431 A)
直流電流／交流＋直流電流Peak<DC Current / AC+DC Current Peak>				
20.00A		±142.5Apeak 50Hz	140.0 A ~ 145.0 A	(142.8 A)
			-145.0 A ~ -140.0 A	(-142.7 A)
600.0A		±1425Apeak 50Hz	1347 A ~ 1503 A	(1434 A)
			-1503 A ~ -1347 A	(-1431 A)
交流電流周波数特性<AC Current Frequency Characteristic>				
20.00A Filter OFF		18.00A 1000Hz	17.54 A ~ 18.46 A	(17.94 A)
600.0A Filter ON		200.0A 10Hz	194.9 A ~ 205.1 A	(199.6 A)
		200.0A 66Hz	196.1 A ~ 203.9 A	(199.2 A)
600.0A Filter OFF		200.0A 10Hz	195.9 A ~ 204.1 A	(199.5 A)
		200.0A 500Hz	195.5 A ~ 204.5 A	(199.6 A)
		200.0A 1000Hz	195.5 A ~ 204.5 A	(199.7 A)
交流＋直流電流周波数特性<AC+DC Current Frequency Characteristic>				
20.00A Filter OFF		18.00A 1000Hz	17.54 A ~ 18.46 A	(17.99 A)
600.0A Filter ON		200.0A 66Hz	195.1 A ~ 204.9 A	(199.2 A)
600.0A Filter OFF		200.0A 1000Hz	195.3 A ~ 204.7 A	(199.6 A)
突入電流／突入電流Peak<Inrush Current / Inrush Current Peak>				
20.00A Inrush rms		18.00A 500Hz	16.97 A ~ 19.03 A	(17.93 A)
600.0A Inrush rms		-540A DC	511.7 A ~ 568.3 A	(540.0 A)
600.0A Inrush peak		-540Apeak DC	-582 A ~ -498 A	(-541 A)
		540Apeak DC	498 A ~ 582 A	(541 A)

備考<Note>

- ・FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>
 - ・形名欄の記載形名は、製品本体表記に基づいています。
- <The model number stated in this document is based on the model number indicated on the instrument.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

SAMPLE
 検査成績表
 <TEST REPORT>

品名<Model Name> (AC/DC クランプメータ<AC/DC CLAMP METER>)
 形名<Model Number> (CM4371-50)
 製造番号<Serial No.> (No.211142205)
 検査年月日<Test Date> (2021-12-04)
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>	
交流電圧<AC Voltage>					
6.000V Filter OFF		0.300V	50Hz	0.295 V ~ 0.305 V	(0.300 V)
		5.400V	50Hz	5.349 V ~ 5.451 V	(5.401 V)
60.00V Filter OFF		3.00V	50Hz	2.95 V ~ 3.05 V	(3.00 V)
		54.00V	50Hz	53.49 V ~ 54.51 V	(54.00 V)
600.0V Filter OFF		30.0V	50Hz	29.5 V ~ 30.5 V	(30.0 V)
		540.0V	50Hz	534.9 V ~ 545.1 V	(540.0 V)
1000V Filter OFF		900V	50Hz	889 V ~ 911 V	(900 V)
直流電圧／交流＋直流電圧<DC Voltage / AC+DC Voltage>					
600.0mV Filter OFF		-540.0mV	DC	-543.2 mV ~ -536.8 mV	(-539.7 mV)
		0.0mV	DC	-0.5 mV ~ 0.5 mV	(0.0 mV)
		540.0mV	DC	536.8 mV ~ 543.2 mV	(540.2 mV)
6.000V Filter OFF		-5.400V	DC	-5.430 V ~ -5.370 V	(-5.400 V)
		0.000V	DC	-0.003 V ~ 0.003 V	(0.000 V)
		5.400V	DC	5.370 V ~ 5.430 V	(5.402 V)
60.00V Filter OFF		-54.00V	DC	-54.30 V ~ -53.70 V	(-54.00 V)
		0.00V	DC	-0.03 V ~ 0.03 V	(0.00 V)
		54.00V	DC	53.70 V ~ 54.30 V	(54.01 V)
600.0V Filter OFF		-540.0V	DC	-543.0 V ~ -537.0 V	(-540.0 V)
		0.0V	DC	-0.3 V ~ 0.3 V	(0.0 V)
		540.0V	DC	537.0 V ~ 543.0 V	(540.1 V)
1000V Filter OFF		-900V	DC	-907 V ~ -893 V	(-900 V)
		900V	DC	893 V ~ 907 V	(900 V)
交流電圧 Peak<AC Voltage Peak>					
6.000V		±10.00Vpeak	50Hz	9.78 V ~ 10.22 V	(9.99 V)
				-10.22 V ~ -9.78 V	(-10.01 V)
60.00V		±100.0Vpeak	50Hz	97.8 V ~ 102.2 V	(100.0 V)
				-102.2 V ~ -97.8 V	(-100.0 V)
600.0V		±1000Vpeak	50Hz	978 V ~ 1022 V	(999 V)
				-1022 V ~ -978 V	(-1000 V)
直流電圧／交流＋直流電圧 Peak<DC Voltage / AC+DC Voltage Peak>					
600.0mV		±1000mVpeak	50Hz	983 mV ~ 1017 mV	(1001 mV)
				-1017 mV ~ -983 mV	(-1000 mV)
6.000V		±10.00Vpeak	50Hz	9.83 V ~ 10.17 V	(10.00 V)
				-10.17 V ~ -9.83 V	(-10.00 V)
60.00V		±100.0Vpeak	50Hz	98.3 V ~ 101.7 V	(100.0 V)
				-101.7 V ~ -98.3 V	(-100.0 V)
600.0V		±1000Vpeak	50Hz	983 V ~ 1017 V	(1000 V)
				-1017 V ~ -983 V	(-1001 V)
交流電圧周波数特性<AC Voltage Frequency Characteristic>					
6.000V Filter ON		5.400V	15Hz	5.287 V ~ 5.513 V	(5.388 V)
			66Hz	5.322 V ~ 5.478 V	(5.389 V)
6.000V Filter OFF		5.400V	15Hz	5.314 V ~ 5.486 V	(5.388 V)
			1000Hz	5.314 V ~ 5.486 V	(5.390 V)
60.00V Filter OFF		54.00V	1000Hz	53.14 V ~ 54.86 V	(53.83 V)
600.0V Filter OFF		540.0V	1000Hz	531.4 V ~ 548.6 V	(538.3 V)
交流＋直流電圧 周波数特性<AC+DC Voltage Frequency Characteristic>					
6.000V Filter ON		5.400V	66Hz	5.306 V ~ 5.494 V	(5.389 V)
6.000V Filter OFF		5.400V	1000Hz	5.306 V ~ 5.494 V	(5.404 V)
60.00V Filter OFF		54.00V	1000Hz	53.06 V ~ 54.94 V	(53.91 V)
600.0V Filter OFF		540.0V	1000Hz	531.2 V ~ 548.8 V	(538.8 V)

検査成績表
<TEST REPORT>

品名<Model Name> (AC/DC クランプメータ<AC/DC CLAMP METER>)
 形名<Model Number> (CM4371-50)
 製造番号<Serial No.> (No.211142205)
 検査年月日<Test Date> (2021-12-04)
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
周波数<Frequency>	9.999Hz	0.600V 9.000Hz	8.988 Hz ~ 9.012 Hz	(9.001 Hz)
	99.99Hz	0.600V 90.00Hz	89.90 Hz ~ 90.10 Hz	(90.01 Hz)
	999.9Hz	0.600V 900.0Hz	899.0 Hz ~ 901.0 Hz	(900.0 Hz)
導通チェック<Continuity Check>	600.0Ω	0.0Ω	-0.5 Ω ~ 0.5 Ω	(0.0 Ω)
		540.0Ω	535.8 Ω ~ 544.2 Ω	(539.8 Ω)
抵抗<Resistance>	600.0Ω	0.0Ω	-0.5 Ω ~ 0.5 Ω	(0.0 Ω)
		540.0Ω	535.8 Ω ~ 544.2 Ω	(539.8 Ω)
	6.000kΩ	0.000kΩ	-0.005 kΩ ~ 0.005 kΩ	(0.000 kΩ)
		5.400kΩ	5.358 kΩ ~ 5.442 kΩ	(5.403 kΩ)
	60.00kΩ	0.00kΩ	-0.05 kΩ ~ 0.05 kΩ	(0.00 kΩ)
		54.00kΩ	53.58 kΩ ~ 54.42 kΩ	(54.01 kΩ)
	600.0kΩ	0.0kΩ	-0.5 kΩ ~ 0.5 kΩ	(0.0 kΩ)
		540.0kΩ	535.8 kΩ ~ 544.2 kΩ	(540.2 kΩ)
	6.000MΩ	0.000MΩ	-0.005 MΩ ~ 0.005 MΩ	(0.000 MΩ)
		5.400MΩ	5.341 MΩ ~ 5.459 MΩ	(5.402 MΩ)
ダイオード<Diode>	1.800V	1.500V	1.485 V ~ 1.515 V	(1.501 V)
静電容量<Capacitance>	1.000μF	0.009μF	0.004 μF ~ 0.014 μF	(0.009 μF)
		0.090μF	0.084 μF ~ 0.096 μF	(0.091 μF)
		0.900μF	0.878 μF ~ 0.922 μF	(0.901 μF)
	10.00μF	9.00μF	8.78 μF ~ 9.22 μF	(8.98 μF)
	100.0μF	90.0μF	87.8 μF ~ 92.2 μF	(90.1 μF)
	1000μF	900μF	878 μF ~ 922 μF	(900 μF)
温度<Temperature>	-40.0~400.0°C	0.0°C	-3.0 °C ~ 3.0 °C	(0.8 °C)
		360.0°C	355.2 °C ~ 364.8 °C	(360.8 °C)

項目 <Item>	レンジ <Range>	入力 <Input>	結果 <Result>
検電<Electric Charge Detection>	HI	0V	(PASS)
		40V 50Hz	(PASS)
	LO	0V	(PASS)
		80V 50Hz	(PASS)
導体位置の影響<Effects Of Conductor Position>	600.0A	200.0A 50Hz	(PASS)