



Certificate of Compliance

Certificate: 80078914

Master Contract: 300342

Project: 80139391

Date Issued: September 20, 2022

Issued To: Hioki E.E. Corporation
81 Koizumi
Ueda, Nagano, 386-1192
Japan

Attention: Inoue Tomoyuki

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Jerry Villamar

PRODUCTS

CLASS - C363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Electrical Measurement and Test Equipment

CLASS - C363186 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

Precision DC Voltmeter, Model DM7275-01, DM7275-02, DM7275-03, DM7276-01, DM7276-02, DM7276-03, Detachable cord connected, rated: 100-240 V, 50/60 Hz, 30 VA.

Notes:

1. The above models are Equipment Class I, Pollution Degree 2, Measurement Category II (Measurement Inputs) and Overvoltage category II (Mains input).
2. Mode of operation: Continuous
3. Environmental Conditions: Normal: 0 °C to 40 °C, 80 % RH or less, non-condensing, 2000 m max



Certificate: 80078914
Project: 80139391

Master Contract: 300342
Date Issued: 2022-09-20

APPLICABLE REQUIREMENTS

CSA Standards:

- | | |
|--|--|
| CAN/CSA C22.2 No. 61010-1-12, UPD1: 2015, UPD2: 2016, AMD1: 2018 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| CAN/CSA-C22.2 No. 61010-2-030-18 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits |

UL Standards:

- | | |
|--|--|
| UL 61010-1, 3 rd edition (2012), AMD1: 2018 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |
| UL Std. No. 61010-2-030:2018 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits |

Notes:

Products certified under Class C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80078914

Master Contract: 300342

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80139391	2022-09-20	Update report: 80008922 to add minor corrections.
80078914	2021-08-03	Update report - 80008922 to add minor design change on non-hazardous circuit and some non-critical components change and minor corrections.