

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
 形名<Model Number> (3200)
 製造番号<Serial No.> (No 110712345)
 検査年月日<Test Date> (2011-12-31)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (25.0 °C, 50 %rh)
 検査周波数<Test Frequency> (50Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
精度<Accuracy>				
DC V	DC 200 mV	DC 190.0 mV	189.2 mV ~ 190.8 mV	(190.0 mV)
	2 V	1.900 V	1.889 V ~ 1.911 V	(1.900 V)
	20 V	19.00 V	18.89 V ~ 19.11 V	(19.00 V)
	200 V	190.0 V	188.9 V ~ 191.1 V	(190.0 V)
	1000 V	1000 V	988 V ~ 1012 V	(1000 V)
AC V	AC 2 V	AC 1.900 V	1.877 V ~ 1.923 V	(1.900 V)
	20 V	19.00 V	18.77 V ~ 19.23 V	(19.00 V)
	200 V	190.0 V	187.7 V ~ 192.3 V	(190.0 V)
	750 V	750 V	739 V ~ 761 V	(750 V)
Ω	200 Ω	190.0 Ω	188.5 Ω ~ 191.5 Ω	(190.0 Ω)
	2 KΩ	1.900 KΩ	1.885 KΩ ~ 1.915 KΩ	(1.900 KΩ)
	20 KΩ	19.00 KΩ	18.85 KΩ ~ 19.15 KΩ	(19.00 KΩ)
	200 KΩ	190.0 KΩ	188.5 KΩ ~ 191.5 KΩ	(190.0 KΩ)
	2000 KΩ	1900 KΩ	1879 KΩ ~ 1921 KΩ	(1900 KΩ)
	20 MΩ	19.00 MΩ	18.60 MΩ ~ 19.40 MΩ	(19.00 MΩ)
DC A	DC 20 μA	DC 19.00 μA	18.79 μA ~ 19.21 μA	(19.00 μA)
	200 μA	190.0 μA	187.9 μA ~ 192.1 μA	(190.0 μA)
	20 mA	19.00 mA	18.79 mA ~ 19.21 mA	(19.00 mA)
	200 mA	190.0 mA	187.9 mA ~ 192.1 mA	(190.0 mA)
	10 A	10.00 A	9.86 A ~ 10.14 A	(10.00 A)
AC A	AC 20 μA	AC 19.00 μA	18.68 μA ~ 19.32 μA	(19.00 μA)
	200 μA	190.0 μA	187.4 μA ~ 192.6 μA	(190.0 μA)
	20 mA	19.00 mA	18.74 mA ~ 19.26 mA	(19.00 mA)
	200 mA	190.0 mA	187.4 mA ~ 192.6 mA	(190.0 mA)
	10 A	10.00 A	9.81 A ~ 10.19 A	(10.00 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()