

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3216)
製造番号<Serial No.> (No. 110912345)
検査年月日<Test Date> (2011-12-31)
<YYYY-MM-DD>
検査条件<Test Conditions> (25.0 °C, 50 %rh)
検査周波数<Test Frequency> (55Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DC V	200 mV	180.0 mV	178.7 mV ~ 181.3 mV	(181.0 mV)
	2 V	1.800 V	1.787 V ~ 1.813 V	(1.800 V)
	20 V	18.00 V	17.84 V ~ 18.16 V	(18.00 V)
	200 V	180.0 V	178.4 V ~ 181.6 V	(180.0 V)
	1000 V	1000 V	986 V ~ 1014 V	(1000 V)
AC V	2 V	1.800 V	1.774 V ~ 1.826 V	(1.800 V)
	20 V	18.00 V	17.74 V ~ 18.26 V	(18.00 V)
	200 V	180.0 V	177.4 V ~ 182.6 V	(180.0 V)
	750 V	750 V	731 V ~ 769 V	(750 V)
Ω	200 Ω	180.0 Ω	178.4 Ω ~ 181.6 Ω	(180.0 Ω)
	2 K Ω	1.800 K Ω	1.784 K Ω ~ 1.816 K Ω	(1.800 K Ω)
	20 K Ω	18.00 K Ω	17.84 K Ω ~ 18.16 K Ω	(18.00 K Ω)
	200 K Ω	180.0 K Ω	178.4 K Ω ~ 181.6 K Ω	(180.0 K Ω)
	2000 K Ω	1800 K Ω	1775 K Ω ~ 1825 K Ω	(1800 K Ω)
	20 M Ω	18.00 M Ω	17.51 M Ω ~ 18.49 M Ω	(18.00 M Ω)
DC A	200 mA	180.0 mA	176.9 mA ~ 183.1 mA	(180.0 mA)
	10 A	10.00 A	9.81 A ~ 10.19 A	(10.00 A)
AC A	200 mA	180.0 mA	175.3 mA ~ 184.7 mA	(180.0 mA)
	10 A	10.00 A	9.70 A ~ 10.30 A	(10.00 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3216-6