

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
形名<Model Number> (3223)
製造番号<Serial No.> (No. 110712345)
検査年月日<Test Date> (2011-12-31)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (25.0 °C, 50 %rh)
検査周波数<Test Frequency> (50Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度					
<Accuracy>	DCV	200 mV	180.00 mV	179.91 mV ~ 180.09 mV	(179.91 mV)
		2000 mV	1800.0 mV	1799.1 mV ~ 1800.9 mV	(1799.1 mV)
		20 V	18.000 V	17.989 V ~ 18.011 V	(17.989 V)
		200 V	180.00 V	179.89 V ~ 180.11 V	(179.89 V)
		1000 V	1000.0 V	999.3 V ~ 1000.7 V	(999.3 V)
	ACV	200 mV	180.00 mV	179.16 mV ~ 180.84 mV	(179.16 mV)
		2000 mV	1800.0 mV	1791.6 mV ~ 1808.4 mV	(1791.6 mV)
		20 V	18.000 V	17.916 V ~ 18.084 V	(17.916 V)
		200 V	180.00 V	179.16 V ~ 180.84 V	(179.16 V)
		750 V	750.0 V	739.5 V ~ 760.5 V	(739.5 V)
	Ω	200 Ω	100.00 Ω	99.91 Ω ~ 100.09 Ω	(99.91 Ω)
		2000 Ω	1000.0 Ω	999.1 Ω ~ 1000.9 Ω	(999.1 Ω)
		20 k Ω	10.000 k Ω	9.991 k Ω ~ 10.009 k Ω	(9.991 k Ω)
		200 k Ω	100.00 k Ω	99.91 k Ω ~ 100.09 k Ω	(99.91 k Ω)
		2000 k Ω	1000.0 k Ω	998.5 k Ω ~ 1001.5 k Ω	(998.5 k Ω)
		20 M Ω	10.000 M Ω	9.940 M Ω ~ 10.060 M Ω	(9.940 M Ω)
	DCA	200 mA	100.00 mA	99.45 mA ~ 100.55 mA	(99.45 mA)
		1000 mA	1000.0 mA	994.5 mA ~ 1005.5 mA	(994.5 mA)
	ACA	200 mA	100.00 mA	98.90 mA ~ 101.10 mA	(98.90 mA)
		1000 mA	1000.0 mA	989.0 mA ~ 1011.0 mA	(989.0 mA)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3223-3