

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (デジタルハイテスタ<DIGITAL HiTESTER>)
 形名<Model Number> (3236)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-03)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (50 Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DC V	300 mV	300.0 mV	298.8 mV ~ 301.2 mV	298.8 mV (
	3 V	3.000 V	2.988 V ~ 3.012 V	2.988 V)
	30 V	30.00 V	29.88 V ~ 30.12 V	29.88 V)
	300 V	300.0 V	298.3 V ~ 301.7 V	298.3 V)
	1000 V	1000 V	992 V ~ 1008 V	992 V)
AC V	3 V	3.000 V	2.966 V ~ 3.034 V	2.966 V (
	30 V	30.00 V	29.66 V ~ 30.34 V	29.66 V)
	300 V	300.0 V	296.6 V ~ 303.4 V	296.6 V)
	750 V	750 V	739 V ~ 761 V	739 V)
Ω	300 Ω	280.0 Ω	278.7 Ω ~ 281.3 Ω	278.7 Ω (
	3 K Ω	2.800 K Ω	2.787 K Ω ~ 2.813 K Ω	2.787 K Ω)
	30 K Ω	28.00 K Ω	27.87 K Ω ~ 28.13 K Ω	27.87 K Ω)
	300 K Ω	280.0 K Ω	278.7 K Ω ~ 281.3 K Ω	278.7 K Ω)
	3000 K Ω	2800 K Ω	2770 K Ω ~ 2830 K Ω	2770 K Ω)
	30 M Ω	28.00 M Ω	27.42 M Ω ~ 28.58 M Ω	27.42 M Ω)
LP Ω	3 K Ω	2.800 K Ω	2.782 K Ω ~ 2.818 K Ω	2.782 K Ω (
	30 K Ω	28.00 K Ω	27.82 K Ω ~ 28.18 K Ω	27.82 K Ω)
	300 K Ω	280.0 K Ω	278.2 K Ω ~ 281.8 K Ω	278.2 K Ω)
	3000 K Ω	2800 K Ω	2768 K Ω ~ 2832 K Ω	2768 K Ω)
	30.0 M Ω	28.00 M Ω	27.40 M Ω ~ 28.60 M Ω	27.40 M Ω)
DC A	30 mA	30.00 mA	29.68 mA ~ 30.32 mA	29.68 mA (
	10 A	10.00 A	9.86 A ~ 10.14 A	9.86 A)
AC A	30 mA	30.00 mA	29.60 mA ~ 30.40 mA	29.60 mA (
	10 A	10.00 A	9.81 A ~ 10.19 A	9.81 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()