

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (クランプ 位相差計<CLAMP ON PHASE HiTESTER>)
形名<Model Number> (3268)
製造番号<Serial No.> (No. 171012315)
検査年月日<Test Date> (2017-04-20)
(YYYY-MM-DD)
検査条件<Test Conditions> (21.3 °C, 55 %rh)
検査周波数<AC Test Frequency> (50Hz)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>		校正値 <Calibration Value>
確度 <Accuracy>	AC mA	AUTO	90.0mA	88.2mA ~	91.8 mA	(89.7 mA)
			45.0mA	43.9mA ~	46.1 mA	(44.9 mA)
			18.0mA	17.3mA ~	18.7 mA	(18.0 mA)
			9.00mA	8.82mA ~	9.18 mA	(8.99 mA)
			4.50mA	4.39mA ~	4.61 mA	(4.50 mA)
			1.80mA	1.73mA ~	1.87 mA	(1.80 mA)
	AC A	AUTO	9.00A	8.82A ~	9.18 A	(9.00 A)
			4.50A	4.39A ~	4.61 A	(4.50 A)
			1.80A	1.73A ~	1.87 A	(1.80 A)
			0.900A	0.882A ~	0.918 A	(0.900 A)
			0.450A	0.439A ~	0.461 A	(0.450 A)
			0.180A	0.173A ~	0.187 A	(0.180 A)
	AC V	AUTO	270.0V	266.8V ~	273.2 V	(270.0 V)
			135.0V	133.2V ~	136.8 V	(135.1 V)
			54.0V	53.0V ~	55.0 V	(54.0 V)
			27.00V	26.68V ~	27.32 V	(27.04 V)
			13.50V	13.32V ~	13.68 V	(13.50 V)
			5.40V	5.30V ~	5.50 V	(5.39 V)

項目 <Item>	ファンクション <Function>	レンジ<Range>		入力<Input>		位相角 <Phase Angle>	許容範囲 <Tolerance>	校正値 <Calibration Value>
位相角 <Phase Angle>	V-A	AUTO	AUTO	100V	90mA	30.0°	27.0° ~33.0°	(30.0 °)
				100V	2mA	30.0°	25.0° ~35.0°	(29.4 °)
				100V	9A	30.0°	27.0° ~33.0°	(30.4 °)
				100V	0.2A	30.0°	27.0° ~33.0°	(29.4 °)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力<Input>		位相角 <Phase Angle>	許容範囲 <Tolerance>	校正値 <Calibration Value>
位相角 <Phase Angle>	V-V	AUTO	100V	Vref	0.0°	358.0° ~ 359.9°	(0.0 °)
						0.0° ~ 2.0°	

機能<Function>		項目<Item>	結果<Result>
DATA OUT<DATA OUT>			(PASS)

備考<Note>
FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()