

検査成績表
TEST REPORT

1/1

品名<Model Name> (クランプオン AC/DC ハイテスタ<CLAMP ON AC/DC HiTESTER>)
 形名<Model Number> (3288-20)
 製造番号<Serial No.> (No. 100399999)
 検査年月日<Test Date> (2010-03-09)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	ファンクション <Function>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度					
<Accuracy> ACA		100.0 A	90.0 A (10Hz)	87.7 A ~ 92.3 A	(90.0 A)
		100.0 A	90.0 A (55Hz)	88.2 A ~ 91.8 A	(90.0 A)
		100.0 A	50.0 A (55Hz)	48.8 A ~ 51.2 A	(50.0 A)
		100.0 A	10.0 A (55Hz)	9.4 A ~ 10.6 A	(10.0 A)
		100.0 A	10.0 A (500Hz)	9.3 A ~ 10.7 A	(10.0 A)
		1000 A	900 A (55Hz)	882 A ~ 918 A	(900 A)
DCA		100.0 A	100.0 A	98.0 A ~ 102.0 A	(100.0 A)
		1000 A	1000 A	980 A ~ 1020 A	(1000 A)
ACV		4.200 V	4.000 V (30Hz)	3.900 V ~ 4.100 V	(4.000 V)
		4.200 V	4.000 V (55Hz)	3.900 V ~ 4.100 V	(4.000 V)
		4.200 V	4.000 V (500Hz)	3.900 V ~ 4.100 V	(4.000 V)
		42.00 V	40.00 V (55Hz)	39.00 V ~ 41.00 V	(40.00 V)
		420.0 V	400.0 V (55Hz)	390.0 V ~ 410.0 V	(400.0 V)
		600 V	600 V (55Hz)	579 V ~ 621 V	(600 V)
DCV		420.0 mV	0.0 mV	-0.4 mV ~ 0.4 mV	(0.0 mV)
		420.0 mV	-400.0 mV	-405.6 mV ~ -394.4 mV	(-400.0 mV)
OHMS		420.0 Ω	0.0 Ω	-0.4 Ω ~ 0.4 Ω	(0.0 Ω)
		420.0 Ω	400.0 Ω	391.6 Ω ~ 408.4 Ω	(400.0 Ω)
		42.00 MΩ	37.00 MΩ	33.26 MΩ ~ 40.74 MΩ	(37.00 MΩ)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.328820-4