

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (クランプオンハイテスタ<CLAMP ON HiTESTER>)
 形名<Model Number> (3291-50)
 製造番号<Serial No.> (No. 100119481)
 検査年月日<Test Date> (2011 - 04- 22)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (22.6 °C, 48 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
確度<Accuracy>				
FILTER ON	1000 A	900 A	55Hz 882 A ~ 918 A	(901 A)
	600 A	540 A	55Hz 531.4 A ~ 548.6 A	(541.0 A)
	60 A	54 A	55Hz 53.14 A ~ 54.86 A	(54.10 A)
FILTER OFF	1000 A	900 A	55Hz 882 A ~ 918 A	(901 A)
	600 A	540 A	55Hz 531.4 A ~ 548.6 A	(540.7 A)
	60 A	54 A	55Hz 53.14 A ~ 54.86 A	(54.07 A)

備考<Nore>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------------	--------------------------	-------------------------