

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (地絡モニタ<GROUND FAULT MONITOR>)
形名<Model Number> (3458)
製造番号<Serial No.> (No.110712345)
検査年月日<Test Date> (2011-12-12)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23°C, 50%rh)

項目 <Item>	入力値 <Input>	許容範囲 <Tolerance>	測定値 <Measurement Value>
--------------	----------------	---------------------	----------------------------

交流電圧精度<AC Voltage Accuracy>

0.0 V 60Hz	0.0 V ~ 0.5 V	(0.0 V)
10.0 V 60Hz	9.4 V ~ 10.6 V	(10.0 V)
30.0 V 60Hz	29.2 V ~ 30.8 V	(30.0 V)
100.0 V 60Hz	98.5 V ~ 101.5 V	(100.0 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3458-2