

**SAMPLE** 検査成績表  
<TEST REPORT>

1/1

品名<Model Name> ( PT )  
 形名<Model Number> ( 9303 )  
 製造番号<Serial No.> ( No.110812345 )  
 検査年月日<Test Date> ( 2011-08-04 )  
 <YYYY-MM-DD>  
 検査条件<Test Conditions> ( 23.0 °C, 50 %rh )  
 検査周波数<Test Frequency> ( 50Hz )

項目 <Item>	電圧比 <Tranceform Retio>	入力電圧 <Input Voltage>	許容範囲 <Tolerance>	標準器指示値 <Indicated Value>
確度 <Accuracy>	40 : 1	400V	9.900V ~ 10.10V	( 10.00 V )
	20 : 1	200V	9.900V ~ 10.10V	( 10.00 V )

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> ( PASS )	検査者<Inspected by> ( )	承認者<Approved by> ( )
----------------------------------	--------------------------	-------------------------