



Certificate of Compliance

Certificate: 80212871 **Master Contract:** 300342
Project: 80212871 **Date Issued:** 2024-07-31
Issued to: Hioki E.E. Corporation
81 Koizumi
Ueda, Nagano 386-1192
Japan
Attention: Inoue Tomoyuki

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by:
Jerry Villamar
Jerry Villamar, Certifier III

PRODUCTS

Class 3631 06 ELECTRICAL MEASUREMENT AND TEST EQUIPMENT - Electrical Measurement and Test Equipment
Class 3631 86 ELECTRICAL MEASUREMENT AND TEST EQUIPMENT - Certified to US Standards

PRECISION BATTERY TESTER

Model(s)	Input Voltage (VAC)	Frequency (Hz)	Power (VA)	Measurement Ratings	Measurement Category
BT6065, BT6075	100 to 240	50/60	40	120VDC	None

Notes:

1. The above model is detachable cord connected, Equipment Class I, Pollution Degree 2, Overvoltage Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: Extended, 0°C to 40°C, max 80% RH (non-condensing), 2000m max



Certificate: 80212871

Master Contract: 300342

Project: 80212871

Date Issued: 2024-07-31

APPLICABLE REQUIREMENTS

CSA C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016, AMD1:2018, UPD3:2023 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016; Update No. 3: June 2023

CSA C22.2 No. 61010-2-030:18 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-030: Particular requirements for equipment having testing or measuring circuits - Second Edition

UL 61010-1 3rd ed (Rev. Jun 6, 2023) - UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Reprint with Revisions Through and Including JUNE 6, 2023

UL 61010-2-030, 2nd Edition (2018) - UL Standard for Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-030: Particular Requirements for Equipment Having Testing or Measuring Circuits



Certificate: 80212871


Project: 80212871

Master Contract: 300342

Date Issued: 2024-07-31

Notes:

Products certified under Class(es) C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80212871

Master Contract: 300342

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80212871	2024-07-31	New model certification for Battery Tester Model: BT6065 and BT6075.