

SAMPLE 検査成績表
<TEST REPORT>

品名<Model Name> (AC/DCカレントセンサ)
<AC/DC CURRENT SENSOR>
形名<Model Number> (CT6875-01)
製造番号<Serial No.> (No. 181212345)
検査年月日<Test Date> (2018-12-12)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)

| 項目 <Item> | 入力 <Input> | 出力基底値 <Fiducial Value> | 許容範囲 <Tolerance> | 校正値 * <Calibration Value> |
|-----------------------------|---------------|---------------------------|---------------------------|------------------------------|
| 1. 振幅確度<Amplitude Accuracy> | | | | |
| | DC 0 A | DC 0 mV | -0.160 mV ~ 0.160 mV | (0.100 mV) |
| | DC 500 A | DC 2 V | 1.999040 V ~ 2.000960 V | (2.000002 V) |
| | DC -500 A | DC -2 V | -2.000960 V ~ -1.999040 V | (-1.999960 V) |
| | AC 500 A | 55Hz AC 2 V | 1.999040 V ~ 2.000960 V | (2.000034 V) |
| | AC 5 A | 55Hz AC 20 mV | 19.832 mV ~ 20.168 mV | (20.000 mV) |
| | AC 5 A | 1kHz AC 20 mV | 19.56 mV ~ 20.44 mV | (20.00 mV) |
| | AC 5 A | 10kHz AC 20 mV | 19.51 mV ~ 20.49 mV | (20.00 mV) |
| | AC 5 A | 100kHz AC 20 mV | 18.40 mV ~ 21.60 mV | (20.01 mV) |
| | AC 5 A | 300kHz AC 20 mV | 17.20 mV ~ 22.80 mV | (19.99 mV) |
| | AC 5 A | 700kHz AC 20 mV | 14.80 mV ~ 25.20 mV | (19.95 mV) |
| | AC 5 A | 1MHz AC 20 mV | 13.00 mV ~ 27.00 mV | (19.99 mV) |

| 項目 <Item> | 入力 <Input> | 許容範囲 <Tolerance> | 校正値 * <Calibration Value> |
|-------------------------|---------------|----------------------|------------------------------|
| 2. 位相確度<Phase Accuracy> | | | |
| | AC 500A 55Hz | -0.10 ° ~ 0.10 ° | (0.01 °) |
| | AC 5A 1kHz | -0.40 ° ~ 0.40 ° | (0.03 °) |
| | AC 5A 10kHz | -1.15 ° ~ 1.15 ° | (-0.24 °) |
| | AC 5A 100kHz | -11.50 ° ~ 11.50 ° | (-2.85 °) |
| | AC 5A 300kHz | -34.50 ° ~ 34.50 ° | (-9.16 °) |
| | AC 5A 700kHz | -80.50 ° ~ 80.50 ° | (-21.53 °) |
| | AC 5A 1MHz | -115.00 ° ~ 115.00 ° | (-29.82 °) |

備考<Note>

1. 位相確度は、入出力間位相差を検査しています。

<Phase accuracy shows phase difference between input and output.>

* FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

| | | |
|----------------------------------|--------------------------|-------------------------|
| 総合判定<Overall Result> (PASS) | 検査者<Inspected By> () | 承認者<Approved By> () |
|----------------------------------|--------------------------|-------------------------|