

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (クランプオンハイテスタ<CLAMP ON HiTESTER>)
 形名<Model Number> (3108)
 製造番号<Serial No.> (No.110712345)
 センサNo.<Sensor No.> (No. 000012)
 検査年月日<Test Date> (2011-12-31)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (55Hz)

項目	レンジ	計器指示値	許容範囲	標準器指示値
<Item>	<Range>	<Indication Value>	<Tolerance>	<Standard>
確度 <Accuracy>	AC 250 A	250 A	243.8 A ~ 256.2 A	(250.0 A)
	100 A	100 A	97.5 A ~ 102.5 A	(100.0 A)
	50 A	50 A	48.75 A ~ 51.25 A	(50.00 A)
	25 A	25 A	24.38 A ~ 25.62 A	(25.00 A)
	10 A	10 A	9.75 A ~ 10.25 A	(10.00 A)
	DC 250 A	250 A	243.8 A ~ 256.2 A	(250.0 A)
	100 A	100 A	97.5 A ~ 102.5 A	(100.0 A)
	50 A	50 A	48.75 A ~ 51.25 A	(50.00 A)
	25 A	25 A	24.38 A ~ 25.62 A	(25.00 A)
	10 A	10 A	9.75 A ~ 10.25 A	(10.00 A)
	+DC 250 A	250 A	+243.8 A ~ +256.2 A	(+ 250.0 A)
	100 A	100 A	+97.5 A ~ +102.5 A	(+ 100.0 A)
	50 A	50 A	+ 48.75 A ~ + 51.25 A	(+ 50.00 A)
	25 A	25 A	+ 24.38 A ~ + 25.62 A	(+ 25.00 A)
	10 A	10 A	+ 9.75 A ~ + 10.25 A	(+ 10.00 A)
	-DC 250 A	250 A	-243.8 A ~ -256.2 A	(- 250.0 A)
	100 A	100 A	-97.5 A ~ -102.5 A	(- 100.0 A)
	50 A	50 A	-48.75 A ~ -51.25 A	(- 50.00 A)
	25 A	25 A	-24.38 A ~ -25.62 A	(- 25.00 A)
	10 A	10 A	-9.75 A ~ -10.25 A	(- 10.00 A)
	AC 500 V	500 V	487.5 V ~ 512.5 V	(500.0 A)
	250 V	250 V	243.8 V ~ 256.2 V	(250.0 A)
	50 V	50 V	48.75 V ~ 51.25 V	(50.00 A)
	25 V	25 V	24.38 V ~ 25.62 V	(25.00 A)
	10 V	10 V	9.75 V ~ 10.25 V	(10.00 A)
	DC 500 V	500 V	487.5 V ~ 512.5 V	(500.0 A)
	250 V	250 V	243.8 V ~ 256.2 V	(250.0 A)
	50 V	50 V	48.75 V ~ 51.25 V	(50.00 A)
	25 V	25 V	24.38 V ~ 25.62 V	(25.00 A)
	10 V	10 V	9.75 V ~ 10.25 V	(10.00 A)
	+DC 500 V	500 V	+487.5 V ~ +512.5 V	(+ 500.0 A)
	250 V	250 V	+243.8 V ~ +256.2 V	(+ 250.0 A)
	50 V	50 V	+48.75 V ~ +51.25 V	(+ 50.00 A)
	25 V	25 V	+24.38 V ~ +25.62 V	(+ 25.00 A)
	10 V	10 V	+9.75 V ~ +10.25 V	(+ 10.00 A)
	-DC 500 V	500 V	-487.5 V ~ -512.5 V	(- 500.0 A)
	250 V	250 V	-243.8 V ~ -256.2 V	(- 250.0 A)
	50 V	50 V	-48.75 V ~ -51.25 V	(- 50.00 A)
	25 V	25 V	-24.38 V ~ -25.62 V	(- 25.00 A)
	10 V	10 V	-9.75 V ~ -10.25 V	(- 10.00 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>

(PASS)

検査者<Inspected by>

()

承認者<Approved by>

()

No.3108-7