

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (クランプオンハイテスタ<CLAMP ON HiTESTER>)
 形名<Model Number> (3109)
 製造番号<Serial No.> (No.110712345)
 センサNo.<Sensor No.> (No. 001111)
 検査年月日<Test Date> (2011-12-31)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (55Hz)

項目	レンジ	計器指示値	許容範囲		標準器指示値
<Item>	<Range>	<Indication value>	<Tolerance>		<Standard>
確 度 <Accuracy>	AC 2500 A	1500 A	1375 A	~	1625 A (1600 A)
	1000 A	1000 A	975 A	~	1025 A (1000 A)
	500 A	500 A	487.5 A	~	512.5 A (500.0 A)
	250 A	250 A	243.8 A	~	256.2 A (250.0 A)
	100 A	100 A	97.5 A	~	102.5 A (100.0 A)
	DC 2500 A	1500 A	1375 A	~	1625 A (1600 A)
	1000 A	1000 A	975 A	~	1025 A (1000 A)
	500 A	500 A	487.5 A	~	512.5 A (500.0 A)
	250 A	250 A	243.8 A	~	256.2 A (250.0 A)
	100 A	100 A	97.5 A	~	102.5 A (100.0 A)
	+DC 2500 A	1500 A	+1375 A	~	+1625 A (+ 1600 A)
	1000 A	1000 A	+975 A	~	+1025 A (+ 1000 A)
	500 A	500 A	+487.5 A	~	+512.5 A (+ 500.0 A)
	250 A	250 A	+243.8 A	~	+256.2 A (+ 250.0 A)
	100 A	100 A	+97.5 A	~	+102.5 A (+ 100.0 A)
	-DC 2500 A	1500 A	-1375 A	~	-1625 A (- 1600 A)
	1000 A	1000 A	-975 A	~	-1025 A (- 1000 A)
	500 A	500 A	-487.5 A	~	-512.5 A (- 500.0 A)
	250 A	250 A	-243.8 A	~	-256.2 A (- 250.0 A)
	100 A	100 A	-97.5 A	~	-102.5 A (- 100.0 A)
	AC 500 V	500 V	487.5 V	~	512.5 V (500.0 A)
	250 V	250 V	243.8 V	~	256.2 V (250.0 A)
	50 V	50 V	48.75 V	~	51.25 V (50.00 A)
	25 V	25 V	24.38 V	~	25.62 V (25.00 A)
	10 V	10 V	9.75 V	~	10.25 V (10.00 A)
	DC 500 V	500 V	487.5 V	~	512.5 V (500.0 A)
	250 V	250 V	243.8 V	~	256.2 V (250.0 A)
	50 V	50 V	48.75 V	~	51.25 V (50.00 A)
	25 V	25 V	24.38 V	~	25.62 V (25.00 A)
	10 V	10 V	9.75 V	~	10.25 V (10.00 A)
	+DC 500 V	500 V	+487.5 V	~	+512.5 V (+ 500.0 A)
	250 V	250 V	+243.8 V	~	+256.2 V (+ 250.0 A)
	50 V	50 V	+48.75 V	~	+51.25 V (+ 50.00 A)
	25 V	25 V	+24.38 V	~	+25.62 V (+ 25.00 A)
	10 V	10 V	+9.75 V	~	+10.25 V (+ 10.00 A)
	-DC 500 V	500 V	-487.5 V	~	-512.5 V (- 500.0 A)
	250 V	250 V	-243.8 V	~	-256.2 V (- 250.0 A)
	50 V	50 V	-48.75 V	~	-51.25 V (- 50.00 A)
	25 V	25 V	-24.38 V	~	-25.62 V (- 25.00 A)
	10 V	10 V	-9.75 V	~	-10.25 V (- 10.00 A)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.3109-7