

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (アナログ接地抵抗計)
<ANALOG EARTH TESTER>

形名<Model Number> (FT3151)

製造番号<Serial No.> (No. 190912345)

検査年月日<Test Date> (2019-09-25)
<YYYY-MM-DD>

検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	校正値 *1 <Calibration Value>
確度<Accuracy>				
接地抵抗測定 3電極法<Earth Resistance Measurement 3-Pole Method>				
	×1 Ω	1	0.75 Ω ~ 1.25 Ω	(1.00 Ω)
		5	4.75 Ω ~ 5.25 Ω	(5.00 Ω)
		10	9.75 Ω ~ 10.25 Ω	(10.00 Ω)
	×10 Ω	1	7.5 Ω ~ 12.5 Ω	(10.0 Ω)
		5	47.5 Ω ~ 52.5 Ω	(50.0 Ω)
		10	97.5 Ω ~ 102.5 Ω	(100.0 Ω)
	×100 Ω	1	75 Ω ~ 125 Ω	(100 Ω)
		5	475 Ω ~ 525 Ω	(500 Ω)
		10	975 Ω ~ 1025 Ω	(1000 Ω)
接地抵抗測定 2電極法<Earth Resistance Measurement 2-Pole Method>				
	×10 Ω	1	7.5 Ω ~ 12.5 Ω	(10.0 Ω)
		5	47.5 Ω ~ 52.5 Ω	(50.0 Ω)
		10	97.5 Ω ~ 102.5 Ω	(100.0 Ω)
	×100 Ω	1	75 Ω ~ 125 Ω	(100 Ω)
		5	475 Ω ~ 525 Ω	(500 Ω)
		10	975 Ω ~ 1025 Ω	(1000 Ω)
地電圧測定<Earth Voltage Measurement> *2				
		30V	29.1 V ~ 30.9 V	(30.0 V)

備考<Note>

*1 FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

*2 検査周波数 60Hz <Test Frequency 60Hz>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()