

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (接地抵抗計<EARTH TESTER>)
 形名<Model Number> (FT6031)^{*1}
 製造番号<Serial No.> (No. 210199731)
 検査年月日<Test Date> (2021-01-19)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (22.2 °C, 37 %RH)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 ^{*2} <Calibration Value>
確度(3電極)<Accuracy (3-Pole)>				
	20 Ω	1 Ω	0.91 Ω ~ 1.09 Ω	(1.00 Ω)
		10 Ω	9.77 Ω ~ 10.23 Ω	(10.01 Ω)
		18 Ω	17.65 Ω ~ 18.35 Ω	(18.03 Ω)
	200 Ω	22 Ω	21.3 Ω ~ 22.7 Ω	(22.0 Ω)
		100 Ω	98.1 Ω ~ 101.9 Ω	(99.6 Ω)
		180 Ω	176.9 Ω ~ 183.1 Ω	(179.3 Ω)
	2000 Ω	220 Ω	213 Ω ~ 227 Ω	(220 Ω)
		1000 Ω	981 Ω ~ 1019 Ω	(1000 Ω)
		1800 Ω	1769 Ω ~ 1831 Ω	(1803 Ω)
確度(2電極)<Accuracy (2-Pole)>				
	200 Ω	22 Ω	18 Ω ~ 26 Ω	(22 Ω)
		100 Ω	95 Ω ~ 105 Ω	(99 Ω)
		180 Ω	174 Ω ~ 186 Ω	(179 Ω)
	2000 Ω	220 Ω	213 Ω ~ 227 Ω	(219 Ω)
		1000 Ω	981 Ω ~ 1019 Ω	(996 Ω)
		1800 Ω	1769 Ω ~ 1831 Ω	(1796 Ω)
確度(地電圧)<Accuracy (Earth Potential)>				
	0V / DC	-0.4 V	~ 0.4 V	(0.0 V)
	28.0V / DC	27.3 V	~ 28.7 V	(28.0 V)
	1.0V / 55Hz	0.2 V	~ 1.8 V	(1.0 V)
	28.0V / 55Hz	26.6 V	~ 29.4 V	(28.0 V)

備考<Note>

・*1 製品本体表記に基づき、形名欄への記載形名を以下のようにしています。<The model number stated in this document correspond to the respective model number indicated on the instrument as listed below:>

FT6031-03: FT6031

FT6031-50: FT6031-50, FT6031-90

・*2 FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()