

SAMPLE

検査成績表
<INSPECTION DATA SHEET>

1/1

品名<Model Name> (電圧ロガー<VOLTAGE LOGGER>)
 形名<Model Number> (LR5041)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Inspection Date> (2011-07-01)
 (<YYYY-MM-DD>)
 検査条件<Inspection Condition> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	検査点 <Inspection Point>	許容範囲 <Tolerance>	表示値 <Indicated Value>
--------------	----------------	---------------------------	---------------------	--------------------------

直流電圧精度<DC Voltage Accuracy>

50 mV	0.00 mV	-0.05 mV ~ 0.05 mV	(0.00 mV)
	40.00 mV	39.75 mV ~ 40.25 mV	(40.00 mV)

総合判定<Judgment> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------	--------------------------	-------------------------