

SAMPLE

検査成績表

<TEST REPORT>

1/2

品名<Model Name> ( デジタルハイテスタ<DIGITAL HITESTER> )  
 形名<Model Number> ( 3251 )  
 製造番号<Serial No.> ( No. 110712345 )  
 検査年月日<Test Date> ( 2011-12-31 )  
 <YYYY-MM-DD>  
 検査条件<Test Conditions> ( 25.0 °C, 50 %rh )  
 交流検査周波数<AC Test Frequency> ( 50Hz )

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DC V	500 mV	450.0 mV	448.5 mV ~ 451.5 mV	( <b>448.5</b> mV )
	5 V	4.500 V	4.485 V ~ 4.515 V	( <b>4.485</b> V )
	50 V	45.00 V	44.85 V ~ 45.15 V	( <b>44.85</b> V )
	500 V	450.0 V	448.5 V ~ 451.5 V	( <b>448.5</b> V )
	1000 V	1000 V	993 V ~ 1007 V	( <b>993</b> V )
DC mV	50 mV	45.00 mV	44.59 mV ~ 45.41 mV	( <b>44.59</b> mV )
	500 mV	450.0 mV	447.6 mV ~ 452.4 mV	( <b>447.6</b> mV )
AC V	500 mV	450.0 mV	442.3 mV ~ 457.7 mV	( <b>442.3</b> mV )
	5 V	4.500 V	4.453 V ~ 4.547 V	( <b>4.453</b> V )
	50 V	45.00 V	44.53 V ~ 45.47 V	( <b>44.53</b> V )
	500 V	450.0 V	445.3 V ~ 454.7 V	( <b>445.3</b> V )
	750 V	750 V	737 V ~ 763 V	( <b>737</b> V )

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
( <b>PASS</b> )	( )	( )

SAMPLE

検査成績表

<TEST REPORT>

製造番号<Serial No.> ( No. 110712345 )  
 検査年月日<Test Date> ( 2011-12-31 )  
 <YYYY-MM-DD>

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated value>
<b>確度&lt;Accuracy&gt;</b>				
DC A	500 $\mu$ A	450.0 $\mu$ A	447.4 $\mu$ A ~ 452.6 $\mu$ A	( <b>447.4</b> $\mu$ A)
	5000 $\mu$ A	4500 $\mu$ A	4476 $\mu$ A ~ 4524 $\mu$ A	( <b>4476</b> $\mu$ A)
	50 mA	45.00 mA	44.74 mA ~ 45.26 mA	( <b>44.74</b> mA)
	500 mA	450.0 mA	447.6 mA ~ 452.4 mA	( <b>447.6</b> mA)
	5 A	4.500 A	4.451 A ~ 4.549 A	( <b>4.451</b> A)
	10 A	9.00 A	8.85 A ~ 9.15 A	( <b>8.85</b> A)
AC A	500 $\mu$ A	450.0 $\mu$ A	440.6 $\mu$ A ~ 459.4 $\mu$ A	( <b>440.6</b> $\mu$ A)
	5000 $\mu$ A	4500 $\mu$ A	4431 $\mu$ A ~ 4569 $\mu$ A	( <b>4431</b> $\mu$ A)
	50 mA	45.00 mA	44.29 mA ~ 45.71 mA	( <b>44.29</b> mA)
	500 mA	450.0 mA	443.1 mA ~ 456.9 mA	( <b>443.1</b> mA)
	5 A	4.500 A	4.429 A ~ 4.571 A	( <b>4.429</b> A)
	10 A	9.00 A	8.80 A ~ 9.20 A	( <b>8.80</b> A)
FREQUENCY	150 Hz	130.00 Hz	129.97 Hz ~ 130.03 Hz	( <b>129.97</b> Hz)
	1500 Hz	1300.0 Hz	1299.7 Hz ~ 1300.3 Hz	( <b>1299.7</b> Hz)
	15 kHz	13.000 kHz	12.997 kHz ~ 13.003 kHz	( <b>12.997</b> kHz)
	150 kHz	130.00 kHz	129.97 kHz ~ 130.03 kHz	( <b>129.97</b> kHz)
	400 kHz	400.0 kHz	399.9 kHz ~ 400.1 kHz	( <b>399.9</b> kHz)
$\Omega$	500 $\Omega$	450.0 $\Omega$	445.1 $\Omega$ ~ 454.9 $\Omega$	( <b>445.1</b> $\Omega$ )
	5 k $\Omega$	4.500 k $\Omega$	4.476 k $\Omega$ ~ 4.524 k $\Omega$	( <b>4.476</b> k $\Omega$ )
	50 k $\Omega$	45.00 k $\Omega$	44.76 k $\Omega$ ~ 45.24 k $\Omega$	( <b>44.76</b> k $\Omega$ )
	500 k $\Omega$	450.0 k $\Omega$	447.6 k $\Omega$ ~ 452.4 k $\Omega$	( <b>447.6</b> k $\Omega$ )
	5 M $\Omega$	4.500 M $\Omega$	4.476 M $\Omega$ ~ 4.524 M $\Omega$	( <b>4.476</b> M $\Omega$ )
	50 M $\Omega$	45.00 M $\Omega$	44.31 M $\Omega$ ~ 45.69 M $\Omega$	( <b>44.31</b> M $\Omega$ )