

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> ( ペンレコーダ<PEN RECORDER> )  
 形名<Model Number> ( PR8111 )  
 製造番号<Serial No.> ( No. 201088111 )  
 検査年月日<Test Date> ( 2020-10-29 )  
 <YYYY-MM-DD>  
 検査条件<Test Conditions> ( 23.7 °C, 54 %rh )

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	結果 <Result>
--------------	----------------	---------------	---------------------	----------------

記録指示試験(フルスケール)

<Full-span Readings>

				CH1
±1mV	2mV	1.990 mV ~	2.010 mV	( PASS )
±2.5mV	5mV	4.975 mV ~	5.025 mV	( PASS )
±5mV	10mV	9.95 mV ~	10.05 mV	( PASS )
±10mV	20mV	19.90 mV ~	20.10 mV	( PASS )
±25mV	50mV	49.75 mV ~	50.25 mV	( PASS )
±50mV	100mV	99.50 mV ~	100.50 mV	( PASS )
±100mV	200mV	199.00 mV ~	201.00 mV	( PASS )
±250mV	500mV	497.50 mV ~	502.50 mV	( PASS )
±500mV	1V	0.995 V ~	1.005 V	( PASS )
±1V	2V	1.990 V ~	2.010 V	( PASS )
±2.5V	5V	4.975 V ~	5.025 V	( PASS )
±5V	10V	9.950 V ~	10.050 V	( PASS )
±10V	20V	19.900 V ~	20.100 V	( PASS )
±25V	50V	49.75 V ~	50.25 V	( PASS )
±50V	100V	99.50 V ~	100.50 V	( PASS )
±100V	200V	199.00 V ~	201.00 V	( PASS )
±250V	250V	247.50 V ~	252.50 V	( PASS )

記録指示試験(マイナス)

<Minus Readings>

±250mV	-250mV	-252.50 mV ~	-247.50 mV	( PASS )
--------	--------	--------------	------------	----------

記録指示試験(ゼロ) <Zero Readings>

±250mV	0mV	-2.50 mV ~	2.50 mV	( PASS )
--------	-----	------------	---------	----------

総合判定<Overall Result> ( PASS )	検査者<Inspected By> ( )	承認者<Approved By> ( )
----------------------------------	--------------------------	-------------------------

SAMPLE

検査成績表

<TEST REPORT>

HIOKI  
HIOKI E. E. CORPORATION

2/2

製造番号<Serial No.> ( No. 201088111 )  
検査年月日<Test Date> ( 2020-10-29 )  
<YYYY-MM-DD>

項目 <Item>	結果 <Result>
スイッチ、ツマミの確認<Switch Operation>	( PASS )
ペンUPLレバー動作確認<PEN Lever Operation>	( PASS )
ペンの脱着<PEN Detachable>	( PASS )
リミッタの検査<Limiter>	( PASS )
記録指示試験(リニアリティ)<Linearity Readings>	( PASS )
スムーズ検査<Smooth Drawing>	( PASS )
片支持落下試験<One-side Drop>	( PASS )
電池駆動試験<Battery Operation>	( PASS )
バッテリーチェック試験<Battery Checker Operation>	( PASS )
チャートスピード検査<Paper-feed Speed>	( PASS )

No.PR8111-5