

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name>	(電源<POWER SUPPLY>)
形名<Model Number>	(3272)
製造番号<Serial No.>	(No.	200412345)
検査年月日<Test Date>	(2020-04-19)
		<YYYY-MM-DD>	
検査条件<Test Conditions>	(23.0 °C, 50 %rh)

項目 <Item>	電源チャネル <Power Supply Receptacles>	許容範囲 <Tolerance>	出力値 <Indicated Value>
出力電圧(負荷±0.3A) <Output Voltage (Load ±0.3A)>	PROBE 1	11.5 V ~ 12.5 V -12.5 V ~ -11.5 V	(12.0 V) (-12.0 V)
	PROBE 2	11.5 V ~ 12.5 V -12.5 V ~ -11.5 V	(12.0 V) (-12.0 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()