

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (アナログユニット<ANALOG UNIT>)
 形名<Model Number> (8932)
 製造番号<Serial No.> (No. 111234567)
 検査年月日<Test Date> (2011-11-03)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 60 %rh)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	測定結果 <Measured Value>
ゼロポジション確度<Zero Position>				
	10 mV/DIV	0 mV	± 1.0 mV	(0.8 mV)
確度<Accuracy>				
	10 mV/DIV	100 mV	± 1.0 mV	(100.8 mV)
	20 mV/DIV	200 mV	± 2.0 mV	(201.6 mV)
	50 mV/DIV	500 mV	± 5.0 mV	(504.0 mV)
	100 mV/DIV	1 V	± 0.010 V	(1.008 V)
	200 mV/DIV	2 V	± 0.020 V	(2.016 V)
	500 mV/DIV	5 V	± 0.050 V	(5.040 V)
	1 V/DIV	10 V	± 0.10 V	(10.08 V)
	2 V/DIV	20 V	± 0.20 V	(20.16 V)
	5 V/DIV	50 V	± 0.50 V	(50.40 V)
	10 V/DIV	100 V	± 1.0 V	(100.8 V)
	20 V/DIV	200 V	± 2.0 V	(201.6 V)
	50 V/DIV	500 V	± 5.0 V	(504.0 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()