

SAMPLE

検査成績表

<INSPECTION DATA SHEET>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (表面抵抗測定用点検治具)
<VERIFICATION FIXTURE FOR SURFACE RESISTANCE MEASUREMENT>
形名<Model Number> (SM9002)
製造番号<Serial No.> (110701234)
検査年月日<Inspection Date> (2011-07-07)
<YYYY-MM-DD>
検査条件<Inspection Condition> (23.0 °C, 50 %rh)

項目 <Item>	許容範囲 <Tolerance>	測定結果 <Measurement>
--------------	---------------------	-----------------------

1.抵抗値検査<Resistance test>

- 1.低抵抗<Low resistance> $4.95 \times 10^5 \Omega \sim 5.05 \times 10^5 \Omega$ ($5.050 \times 10^5 \Omega$)
-2.高抵抗<High resistance> $0.95 \times 10^{12} \Omega \sim 1.05 \times 10^{12} \Omega$ ($1.050 \times 10^{12} \Omega$)

総合判定<Judgment>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

SM9002-1