

検査成績表

<TEST REPORT>

品名<Model Name> (電流ユニット<CURRENT UNIT>)
 形名<Model Number> (8971)
 製造番号<Serial No.> (No. 110724125)
 検査年月日<Test Date> (2011-07-19)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (25.8 °C, 65 %rh)

項目 <Item>	レンジ <Range>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>	
1. 測定精度<Measurement Accuracy>				CH1 CH2	
20A/2V	100 mA/DIV	0 A	-0.013 A ~ 0.013 A	(-0.001 A) (0.000 A)	
	100 mA/DIV	2 A	1.987 A ~ 2.013 A	(2.001 A) (2.001 A)	
	100 mA/DIV	-2 A	-2.013 A ~ -1.987 A	(-2.001 A) (-2.000 A)	
	200 mA/DIV	4 A	3.974 A ~ 4.026 A	(3.996 A) (3.998 A)	
	500 mA/DIV	10 A	9.935 A ~ 10.065 A	(10.000 A) (10.005 A)	
	1 A/DIV	20 A	19.87 A ~ 20.13 A	(19.99 A) (19.99 A)	
	2 A/DIV	40 A	39.74 A ~ 40.26 A	(40.00 A) (39.98 A)	
	5 A/DIV	100 A	99.35 A ~ 100.65 A	(100.00 A) (100.00 A)	
	500A/2V	2 A/DIV	0 A	-0.26 A ~ 0.26 A	(0.00 A) (0.00 A)
		2 A/DIV	40 A	39.74 A ~ 40.26 A	(40.02 A) (40.02 A)
		2 A/DIV	-40 A	-40.26 A ~ -39.74 A	(-39.98 A) (-40.00 A)
		5 A/DIV	100 A	99.35 A ~ 100.65 A	(99.90 A) (99.90 A)
		10 A/DIV	200 A	198.7 A ~ 201.3 A	(200.0 A) (200.0 A)
		20 A/DIV	400 A	397.4 A ~ 402.6 A	(400.0 A) (399.8 A)
50 A/DIV	1000 A	993.5 A ~ 1006.5 A	(1000.0 A) (999.0 A)		
100 A/DIV	2000 A	1987 A ~ 2013 A	(1999 A) (2000 A)		
2. RMS精度<RMS Accuracy>					
20A/2V	100 mA/DIV	0 A	-0.020 A ~ 0.020 A	(-0.001 A) (0.000 A)	
	100 mA/DIV	2 A	1.980 A ~ 2.020 A	(2.003 A) (2.002 A)	
	100 mA/DIV	-2 A	1.980 A ~ 2.020 A	(2.003 A) (2.003 A)	
	200 mA/DIV	4 A	3.960 A ~ 4.040 A	(4.000 A) (4.002 A)	
	500 mA/DIV	10 A	9.900 A ~ 10.100 A	(10.000 A) (10.000 A)	
	1 A/DIV	20 A	19.80 A ~ 20.20 A	(19.99 A) (19.99 A)	
	2 A/DIV	40 A	39.60 A ~ 40.40 A	(40.02 A) (40.02 A)	
	5 A/DIV	100 A	99.00 A ~ 101.00 A	(99.95 A) (100.05 A)	
	500A/2V	2 A/DIV	0 A	-0.40 A ~ 0.40 A	(-0.02 A) (0.00 A)
		2 A/DIV	40 A	39.60 A ~ 40.40 A	(40.06 A) (40.02 A)
		2 A/DIV	-40 A	39.60 A ~ 40.40 A	(40.06 A) (40.06 A)
		5 A/DIV	100 A	99.00 A ~ 101.00 A	(100.00 A) (100.05 A)
		10 A/DIV	200 A	198.0 A ~ 202.0 A	(200.0 A) (200.1 A)
		20 A/DIV	400 A	396.0 A ~ 404.0 A	(399.8 A) (399.8 A)
50 A/DIV	1000 A	990.0 A ~ 1010.0 A	(1000.5 A) (1000.0 A)		
100 A/DIV	2000 A	1980 A ~ 2020 A	(2001 A) (2001 A)		

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------------	--------------------------	-------------------------