

Sample

## 検査成績表

&lt;TEST REPORT&gt;

1/2

品名<Model Name> ( パワーハイテスタ<POWER HiTESTER> )  
形名<Model Number> ( 3332 )  
製造番号<Serial No.> ( No. 210912345 )  
検査年月日<Test Date> ( 2021-10-01 )  
<YYYY-MM-DD>  
検査条件<Test Conditions> ( 23.7 °C, 42 %rh )  
検査周波数<Test Frequency> ( 55 Hz )

項目 <Item>	レクティファイア／レスポンス <RECT> <RESP>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
精度<Accuracy>					
RMS	FAST	600 V	600.00 V	598.80 V ~ 601.20 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			150.00 V	149.55 V ~ 150.45 V	( V )
			30.00 V	29.67 V ~ 30.33 V	( V )
			0.00 V	-0.30 V ~ 0.30 V	( V )
		150 V	150.00 V	149.70 V ~ 150.30 V	( V )
		60 V	60.000 V	59.880 V ~ 60.120 V	( V )
		30 V	30.000 V	29.940 V ~ 30.060 V	( V )
		15 V	15.000 V	14.970 V ~ 15.030 V	( V )
		300 V	300.00 V	299.40 V ~ 300.60 V	( V )
	SLOW	MEAN	300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
		MEAN+FILT	300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
		RMS	300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
			300.00 V	299.40 V ~ 300.60 V	( V )
RMS	FAST	50 A	20.000 A	19.930 A ~ 20.070 A	( A )
			20.000 A	19.960 A ~ 20.040 A	( A )
			10.000 A	9.980 A ~ 10.020 A	( A )
			5.0000 A	4.9900 A ~ 5.0100 A	( A )
			2.5000 A	2.4925 A ~ 2.5075 A	( A )
		0.5000 A	0.4945 A	0.4945 A ~ 0.5055 A	( A )
			0.0000 A	-0.0050 A ~ 0.0050 A	( A )
			2.0000 A	1.9960 A ~ 2.0040 A	( A )
			1.0000 A	0.9980 A ~ 1.0020 A	( A )
			0.5000 A	499.00 mA ~ 501.00 mA	( mA )
	SLOW	0.2 A	200.00 mA	199.60 mA ~ 200.40 mA	( mA )
			100.00 mA	99.80 mA ~ 100.20 mA	( mA )
			50.000 mA	49.900 mA ~ 50.100 mA	( mA )
			20.000 mA	19.960 mA ~ 20.040 mA	( mA )
			10.000 mA	9.980 mA ~ 10.020 mA	( mA )
		5 mA	5.0000 mA	4.9900 mA ~ 5.0100 mA	( mA )
			2.0000 mA	1.9960 mA ~ 2.0040 mA	( mA )
			1.0000 mA	0.9980 mA ~ 1.0020 mA	( mA )
			5.0000 A	4.9900 A ~ 5.0100 A	( A )
			5.0000 A	4.9900 A ~ 5.0100 A	( A )
<PEAK>	FAST	5 A	7.071A (PEAK)	6.734 A ~ 7.408 A	( A )

総合判定&lt;Overall Result&gt;

( )

検査者&lt;Inspected by&gt;

( S. Norinchi )

承認者&lt;Approved by&gt;

( Y. Sumohara )

No. 3332-5A

# 検査成績表

<TEST REPORT>

2/2

製造番号<Serial No.> ( No. 210912345 )  
検査年月日<Test Date> ( 2021-10-01 )  
<YYYY-MM-DD>

項目 ＜Item＞	レクティファイ ／レスポンス		レンジ ＜Range＞		入力 ＜Input＞			許容範囲 ＜Tolerance＞				表示値 ＜Indicated Value＞	
	＜RECT＞	＜RESP＞	電圧 ＜U＞	電流 ＜I＞	電圧 ＜U＞	電流 ＜I＞	電力 ＜P＞						
確度＜Accuracy＞													
RMS	FAST	150V	5A	150V	5A	750W	748.50	W	～	751.50	W	(	W)
				150V	2.5A	375W	373.88	W	～	376.12	W	(	W)
				150V	0.5A	75W	74.18	W	～	75.82	W	(	W)
				0V	0A	0W	－0.75	W	～	0.75	W	(	W)
	SLOW	150V	5A	150V	5A	750W	748.50	W	～	751.50	W	(	W)

項目 <Item>	レンジ <Range> 電圧 <U> 周波数 <FREQ>	入力 <Input> 電圧 <U> 周波数 <FREQ>	許容範囲 <Tolerance>	表示値 <Indicated Value>
周波数 <Frequency>	300V 500Hz 5A 500Hz	300V 55.000Hz 5A 55.000Hz	54.944 Hz ~ 55.056 Hz 54.944 Hz ~ 55.056 Hz	( Hz) ( Hz)

## 備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

No.3332-5A