

Product Specification Changes and Software Version Upgrade for the Clamp-on Power Logger PW3360 Series due to Component Changes

Due to the discontinuation of certain components used in the Clamp-On Power Logger PW3360 series, we have made changes to components and circuit boards.

As a result, there are changes to the software version and continuous battery operation time, etc.

1. Products with these changes:

Products produced or replaced with board replacement after mid-September 2024

2. Applicable products and details:

Clamp-on Power Logger PW3360 Series

PW3360-10, PW3360-11, PW3360-20, PW3360-21, PW3360-30, PW3360-31, PW3360-91

	Before	After
Software version	Until V3.22	V5.22
Continuous battery operation time (When using a battery pack, continuous, backlight OFF)	Approx. 8 hours	Approx. 6 hours
Software start-up time (From when the power is switched on to when the opening screen is displayed)	Approx. 1 second	Approx. 3 seconds
Maximum rated power (when using the Battery Pack 9459)	3 VA	3 VA (no change)
Enclosure temperature	-	About 10°C higher than before the change