

Product Specification Changes and Software Version Upgrade for the Clamp-on Power Logger PW3365 Series due to Component Changes

Due to the discontinuation of certain components used in the Clamp-On Power Logger PW3365 series, we have made changes to components and circuit boards.

As a result, there are changes to the software version and continuous battery operation time, etc.

1. Products with these changes:

Products produced or replaced with board replacement after mid-September 2024

2. Applicable products and details:

Clamp-on Power Logger PW3365 Series

PW3365-10,PW3365-20,PW3365-30

	Before	After
Software version	Until V2.11	V5.11
Continuous battery operation time (When using a battery pack, continuous, backlight OFF, when using four Safety Voltage Sensor PW9020)	Approx. 5 hours	Approx. 3 hours
Software start-up time (From when the power is switched on to when the opening screen is displayed)	Approx. 1 second	Approx. 3 seconds
Maximum rated power (when using the Battery Pack 9459)	4 VA	5 VA
Enclosure temperature	-	About 5°C higher than before the change