

Request for the re-adjustment of INSULATOIN RESISTANCE TESTER

Dear valued customers,

Thank you very much for your continued patronage.

We have found a possibility that the measurement value may become the out of the accuracy on models IR4051-10, IR4051-11, IR4052-50, IR4052-51, IR4052-91, IR4056-20, IE4056-21, IR4056-36, and IR4057-50.

We would like to re-adjust the subjected products because this may affect your business operations. Sorry for your inconvenience, but we appreciate your kind cooperation in advance.

1. Subjected products

INSULATION TESTER

IR4051-10, IR4051-11, IR4052-50, IR4052-51, IR4052-91, IR4056-20, IE4056-21, IR4056-36, IR4057-50

Subjected period: Shipped from our factory from March, 2025 to January, 2026

Subjected serial No. Please refer to the list in below (98 units).

2. Details of failure

The insulation resistance measurement may become out of the accuracy when measuring specific ranges in below.

Rated measurement voltage	Resistance may become out of the accuracy specifications
50 V	0.160 MΩ to 1.20 MΩ
125 V	0.400 MΩ to 3.00 MΩ
250 V	0.800 MΩ to 6.00 MΩ
500 V	1.60 MΩ to 12.0 MΩ
1000 V	3.20 MΩ to 24.0 MΩ

3. Request for re-adjustment

We would like to re-adjust the subjected products at free. Please be consulted with our nearest sales office for details.

List of serial number of subjected products

(Serial number is common to all models)

IR4051 (-10,-11)		IR4052 (-50,-51,-91)	IR4056 (-20,-21,-36)		IR4057-50
250405058	251098945	250328945	250412575	250609255	250412672
250415556	251099160	250524787	250412584	250609299	250332332
250416978	251099162	250524798	250412596	250612841	250429342
250416986	251106779	250524802	250414275	250720286	250429347
250416991	251204148	250524804	250414283	250725783	250502486
250416999	251209737	250526516	250414287	250730376	250503475
250418732		250605054	250414288	250824650	250503476
250419832		250605056	250417068	250910326	250504851
250421223		250605070	250417074	251098893	250504852
250421254		250610174	250417075	251098915	
250421259		250610179	250421303	251104080	
250424540			250421338	251104083	
250425138			250425233	251104111	
250425150			250428336	251121984	
250428244			250428348	251199766	
250428292			250429269	251199768	
250428295			250515732	251226331	
250503416			250517298	251229859	
250504772			250523755	251229867	
250505898			250603941	260107598	
250505928			250605137	260109010	
250507067			250607754	260115127	
250525643			250609235	260119104	
250609199			250609252	260120801	