

SAMPLE

検査成績表
(TEST REPORT)

品名<Model Name> (パワーメータ<POWER METER>)
 形名<Model Number> (PW3336)
 製造番号<Serial No.> (No. 130501925)
 検査年月日<Test Date> (2021-05-25)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (24.1 °C, 49 %rh)

1. 測定精度<Measurement Accuracy>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1		
			<Volt>	<Freq>		CH1	CH2	
-1. 電圧<Voltage>								
AC+DC		1000V	1kV	55Hz	0.9985 kV ~ 1.0015 kV	(1.0000 kV)	(1.0000 kV)	
			0V	-	0.0000 kV ~ 0.0000 kV	(0.0000 kV)	(0.0000 kV)	
			600V	55Hz	599.10 V ~ 600.90 V	(600.01 V)	(600.01 V)	
		600V	300V	55Hz	299.55 V ~ 300.45 V	(300.00 V)	(299.99 V)	
			150V	55Hz	149.78 V ~ 150.22 V	(150.00 V)	(150.00 V)	
		100V	100V	55Hz	99.85 V ~ 100.15 V	(100.00 V)	(100.00 V)	
			75V	55Hz	74.89 V ~ 75.11 V	(75.00 V)	(75.00 V)	
			1.5V	55Hz	1.43 V ~ 1.57 V	(1.50 V)	(1.50 V)	
		0V	0V	-	0.00 V ~ 0.00 V	(0.00 V)	(0.00 V)	
			60V	55Hz	59.910 V ~ 60.090 V	(60.000 V)	(60.000 V)	
			30V	55Hz	29.955 V ~ 30.045 V	(30.000 V)	(30.000 V)	
		15V	15V	55Hz	14.978 V ~ 15.022 V	(15.000 V)	(15.000 V)	
			0V	-	0.000 V ~ 0.000 V	(0.000 V)	(0.000 V)	
			DC	1000V	+1kV		0.9980 kV ~ 1.0020 kV	(1.0000 kV)
		0V			0.0000 kV ~ 0.0000 kV	(0.0000 kV)	(0.0000 kV)	
-1kV		-1.0020 kV ~ -0.9980 kV	(-1.0000 kV)		(-1.0000 kV)			
600V	+600V		598.80 V ~ 601.20 V		(600.07 V)	(600.04 V)		
	0V		0.00 V ~ 0.00 V		(0.00 V)	(0.00 V)		
300V	+300V		299.40 V ~ 300.60 V		(300.03 V)	(300.02 V)		
	0V		0.00 V ~ 0.00 V		(0.00 V)	(0.00 V)		
150V	+150V		149.70 V ~ 150.30 V		(150.01 V)	(150.01 V)		
	0V		0.00 V ~ 0.00 V		(0.00 V)	(0.00 V)		
60V	+60V		59.880 V ~ 60.120 V		(60.004 V)	(59.999 V)		
	0V		0.000 V ~ 0.000 V		(0.000 V)	(0.000 V)		
30V	+30V		29.940 V ~ 30.060 V		(30.001 V)	(30.002 V)		
	0V		0.000 V ~ 0.000 V		(0.000 V)	(0.000 V)		
15V	+15V		14.970 V ~ 15.030 V		(15.000 V)	(15.000 V)		
	+0.15V		0.135 V ~ 0.165 V		(0.150 V)	(0.151 V)		
	0V		0.000 V ~ 0.000 V	(0.000 V)	(0.000 V)			
		-15V		-15.030 V ~ -14.970 V	(-15.001 V)	(-14.999 V)		

備考<Note>

・特に設定の明記のない場合、以下の設定で検査しています。

結線: 1P2W, VT比: 1, CT比: 1, ゼロクロスフィルタ: 500Hz

<Inspection settings are as follows except noted.

Wiring: 1P2W, VT Ratio: 1, CT Ratio: 1, Filter(Zero Cross): 500Hz>

*1. FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

SAMPLE

検査成績表
<TEST REPORT>

製造番号<Serial No.>
検査年月日<Test Date>

(No. 130501925)
(2021-05-25)
<YYYY-MM-DD>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1		
			<Curr>	<Freq>		CH1	CH2	
-2. 電流<Current>								
	AC+DC	50A	50A	55Hz	49.925 A ~ 50.075 A	(50.000 A)	(49.996 A)	
			0A	-	0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	
		20A	20A	55Hz	19.970 A ~ 20.030 A	(20.000 A)	(20.000 A)	
		10A	10A	55Hz	9.985 A ~ 10.015 A	(10.008 A)	(10.008 A)	
		5A	5A	55Hz	4.9925 A ~ 5.0075 A	(5.0000 A)	(5.0000 A)	
			2.5A	55Hz	2.4963 A ~ 2.5037 A	(2.4999 A)	(2.5000 A)	
			0.5A	55Hz	0.4970 A ~ 0.5030 A	(0.5001 A)	(0.5002 A)	
			0.05A	55Hz	0.0475 A ~ 0.0525 A	(0.0501 A)	(0.0501 A)	
			0A	-	0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	
		2A	2A	55Hz	1.9970 A ~ 2.0030 A	(2.0000 A)	(2.0000 A)	
		1A	1A	55Hz	0.9985 A ~ 1.0015 A	(1.0008 A)	(1.0008 A)	
			0.01A	55Hz	0.0095 A ~ 0.0105 A	(0.0101 A)	(0.0101 A)	
		500mA	500mA	55Hz	499.25 mA ~ 500.75 mA	(500.00 mA)	(500.00 mA)	
		200mA	200mA	55Hz	199.70 mA ~ 200.30 mA	(200.04 mA)	(200.04 mA)	
			2mA	55Hz	1.90 mA ~ 2.10 mA	(2.02 mA)	(2.02 mA)	
			0A	-	0.00 mA ~ 0.00 mA	(0.00 mA)	(0.00 mA)	
	DC	50A	+50A		49.899 A ~ 50.101 A	(49.998 A)	(49.998 A)	
			0A		0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	
			-50A		-50.101 A ~ -49.899 A	(-49.998 A)	(-49.998 A)	
		20A	+20A		19.959 A ~ 20.041 A	(19.993 A)	(19.997 A)	
			0A		0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	
		10A	+10A		9.979 A ~ 10.021 A	(10.000 A)	(10.001 A)	
			0A		0.000 A ~ 0.000 A	(0.000 A)	(0.000 A)	
		5A	+5A		4.9890 A ~ 5.0110 A	(5.0002 A)	(5.0011 A)	
			0A		0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	
		2A	+2A		1.9950 A ~ 2.0050 A	(2.0002 A)	(2.0002 A)	
			0A		0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	
		1A	+1A		0.9970 A ~ 1.0030 A	(1.0001 A)	(1.0002 A)	
			0A		0.0000 A ~ 0.0000 A	(0.0000 A)	(0.0000 A)	
			-1A		-1.0030 A ~ -0.9970 A	(-1.0002 A)	(-1.0002 A)	
		500mA	+500mA		498.00 mA ~ 502.00 mA	(500.06 mA)	(500.07 mA)	
			0A		0.00 mA ~ 0.00 mA	(0.00 mA)	(0.00 mA)	
		200mA	+200mA		198.60 mA ~ 201.40 mA	(199.99 mA)	(199.97 mA)	
			+2mA		0.80 mA ~ 3.20 mA	(2.00 mA)	(1.93 mA)	
			0A		0.00 mA ~ 0.00 mA	(0.00 mA)	(0.00 mA)	
			-200mA		-201.40 mA ~ -198.60 mA	(-200.02 mA)	(-200.06 mA)	

SAMPLE

検査成績表
<TEST REPORT>

製造番号<Serial No.>
検査年月日<Test Date>

(No. 130501925)
(2021-05-25)
<YYYY-MM-DD>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input> <Volt/Curr/Pow/Freq/PF>	許容範囲 <Tolerance>	校正値<Calibration Value> *1	
					CH1	CH2
-3. 有効電力 力率 1<Power PF=1>						
	AC+DC	150V /5A	150V, 5A, 750W, 55Hz, PF=1	748.88 W ~ 751.12 W (749.76 W) (749.75 W)
			150V, 2.5A, 375W, 55Hz, PF=1	374.44 W ~ 375.56 W (374.91 W) (374.93 W)
			150V, 0.5A, 75W, 55Hz, PF=1	74.55 W ~ 75.45 W (74.99 W) (75.02 W)
			150V,0.05A,7.5W, 55Hz, PF=1	7.12 W ~ 7.88 W (7.51 W) (7.50 W)
			150V(55Hz), 0A, 0W, -, -	-0.37 W ~ 0.37 W (0.00 W) (0.00 W)
		1000V /50A	1000V, 50A, 50kW, 55Hz, PF=1	49.925 kW ~ 50.075 kW (50.012 kW) (50.009 kW)
		1000V /200mA	1000V, 200mA, 200W, 55Hz, PF=1	199.70 W ~ 200.30 W (200.04 W) (200.04 W)
		15V /50A	15V, 50A, 750W, 55Hz, PF=1	748.88 W ~ 751.12 W (750.21 W) (750.15 W)
		15V /200mA	15V, 200mA, 3W, 55Hz, PF=1	2.9955 W ~ 3.0045 W (3.0000 W) (3.0000 W)
			15V(55Hz), 0A, 0W, -, -	-0.0015 W ~ 0.0015 W (0.0000 W) (0.0000 W)
	DC	15V /200mA	+15V, +200mA, +3W	2.9790 W ~ 3.0210 W (3.0000 W) (2.9995 W)
			-15V, -200mA, +3W	2.9790 W ~ 3.0210 W (3.0004 W) (3.0007 W)
			+15V, +2mA, +0.03W	0.0120 W ~ 0.0480 W (0.0299 W) (0.0290 W)
			0V, 0A, 0W	-0.0030 W ~ 0.0030 W (0.0000 W) (0.0000 W)
		1000V /50A	+1000V, +50A, +50kW	49.899 kW ~ 50.101 kW (49.987 kW) (49.989 kW)
			-1000V, -50A, +50kW	49.899 kW ~ 50.101 kW (50.003 kW) (49.992 kW)

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input> <Volt/Curr/Pow/Freq/PF>	許容範囲 <Tolerance>	校正値<Calibration Value> *1	
					CH1	CH2
-4. 有効電力 力率 0<Power PF=0>						
	AC+DC	150V /5A	150V, 5A, 0W, 55Hz, LEAD 0	-0.75 W ~ 0.75 W (0.03 W) (0.02 W)
			150V, 5A, 0W, 55Hz, LAG 0	-0.75 W ~ 0.75 W (-0.03 W) (-0.03 W)
		15V /20A	15V, 20A, 0W, 55Hz, LEAD 0	-0.30 W ~ 0.30 W (0.01 W) (0.01 W)
			15V, 20A, 0W, 55Hz, LAG 0	-0.30 W ~ 0.30 W (-0.01 W) (-0.01 W)

SAMPLE

検査成績表
<TEST REPORT>

製造番号<Serial No.> (No. 130501925)
 検査年月日<Test Date> (2021-05-25)
 <YYYY-MM-DD>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1					
			<Curr>	<Freq>		CH1	CH2				
-5. 外部センサ TYPE 1 CT比:10<External Sensor TYPE 1 CT Ratio:10>											
AC+DC	500A	500A	500A	55Hz	498.50 A ~ 501.50 A (500.02 A)	(500.02 A)			
			0A	-	0.00 A ~ 0.00 A (0.00 A)	(0.00 A)			
			200A	55Hz	199.40 A ~ 200.60 A (199.99 A)	(199.99 A)			
			100A	55Hz	99.70 A ~ 100.30 A (100.01 A)	(100.01 A)			
			50A	55Hz	49.85 A ~ 50.15 A (50.01 A)	(50.01 A)			
			10A	55Hz	9.88 A ~ 10.12 A (9.99 A)	(10.00 A)			
			1A	55Hz	0.90 A ~ 1.10 A (0.99 A)	(0.99 A)			
			0A	-	0.00 A ~ 0.00 A (0.00 A)	(0.00 A)			
			DC	500A	500A	500A		496.00 A ~ 504.00 A (500.09 A)	(500.18 A)
						0A		0.00 A ~ 0.00 A (0.00 A)	(0.00 A)
						200A		198.40 A ~ 201.60 A (200.13 A)	(200.14 A)
						0A		0.00 A ~ 0.00 A (0.00 A)	(0.00 A)
100A		99.20 A ~ 100.80 A (100.12 A)	(100.14 A)			
1A		0.40 A ~ 1.60 A (1.11 A)	(1.11 A)			
0A		0.00 A ~ 0.00 A (0.00 A)	(0.00 A)						
		-100A		-100.80 A ~ -99.20 A (-99.85 A)	(-99.85 A)				

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1					
			<Curr>	<Freq>		CH1	CH2				
-6. 外部センサ TYPE 2 CT比:1<External Sensor TYPE 2 CT Ratio:1>											
AC+DC	50A	50A	50A	55Hz	49.850 A ~ 50.150 A (50.008 A)	(50.009 A)			
			0A	-	0.000 A ~ 0.000 A (0.000 A)	(0.000 A)			
			20A	55Hz	19.940 A ~ 20.060 A (19.998 A)	(19.998 A)			
			10A	55Hz	9.970 A ~ 10.030 A (10.000 A)	(10.000 A)			
			5A	55Hz	4.985 A ~ 5.015 A (5.001 A)	(5.001 A)			
			1A	55Hz	0.988 A ~ 1.012 A (1.000 A)	(1.001 A)			
			0.1A	55Hz	0.090 A ~ 0.110 A (0.100 A)	(0.100 A)			
			0A	-	0.000 A ~ 0.000 A (0.000 A)	(0.000 A)			
			DC	50A	50A	50A		49.600 A ~ 50.400 A (49.993 A)	(49.999 A)
						0A		0.000 A ~ 0.000 A (0.000 A)	(0.000 A)
						20A		19.840 A ~ 20.160 A (20.000 A)	(20.003 A)
						0A		0.000 A ~ 0.000 A (0.000 A)	(0.000 A)
10A		9.920 A ~ 10.080 A (10.004 A)	(10.005 A)			
0.1A		0.040 A ~ 0.160 A (0.103 A)	(0.103 A)			
0A		0.000 A ~ 0.000 A (0.000 A)	(0.000 A)						
		-10A		-10.080 A ~ -9.920 A (-9.997 A)	(-9.996 A)				

SAMPLE

検査成績表
(TEST REPORT)

製造番号<Serial No.>
検査年月日<Test Date>

(No. 130501925)
(2021-05-25)
<YYYY-MM-DD>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input> <Volt/Curr/Pow/Freq/PF>	許容範囲 <Tolerance>	校正値<Calibration Value> *1		
					CH1	CH2	
-7. 有効電力 力率 1 外部センサ TYPE 1 CT比:10<Power PF=1 External Sensor TYPE 1 CT Ratio:10>							
AC+DC	レクティ ファイア	150V /200A	150V, 200A, 30kW, 55Hz, PF=1	29.910 kW ~ 30.090 kW (30.000 kW) (30.000 kW)	
			150V, 100A, 15kW, 55Hz, PF=1	14.955 kW ~ 15.045 kW (15.004 kW) (15.005 kW)	
			150V, 20A, 3kW, 55Hz, PF=1	2.964 kW ~ 3.036 kW (3.000 kW) (3.001 kW)	
			150V, 2A, 0.3kW, 55Hz, PF=1	0.270 kW ~ 0.330 kW (0.303 kW) (0.303 kW)	
			150V(55Hz), 0A, 0W, -, -	-0.030 kW ~ 0.030 kW (0.000 kW) (0.000 kW)	
			1000V /500A	1000V, 500A, 500kW, 55Hz, PF=1	498.50 kW ~ 501.50 kW (500.12 kW) (500.12 kW)
			1000V /100A	1000V, 100A, 100kW, 55Hz, PF=1	99.70 kW ~ 100.30 kW (100.03 kW) (100.03 kW)
			15V /500A	15V, 500A, 7.5kW, 55Hz, PF=1	7.4775 kW ~ 7.5225 kW (7.5022 kW) (7.5021 kW)
			15V /100A	15V, 100A, 1.5kW, 55Hz, PF=1	1.4955 kW ~ 1.5045 kW (1.5003 kW) (1.5005 kW)
				15V(55Hz), 0A, 0W, -, -	-0.0015 kW ~ 0.0015 kW (0.0000 kW) (0.0000 kW)
DC	レクティ ファイア	15V /100A	+15V, +100A, +1.5kW	1.4880 kW ~ 1.5120 kW (1.5018 kW) (1.5022 kW)	
			-15V, -100A, +1.5kW	1.4880 kW ~ 1.5120 kW (1.4981 kW) (1.4978 kW)	
			+15V, +1A, +0.015kW	0.0060 kW ~ 0.0240 kW (0.0167 kW) (0.0167 kW)	
			0V, 0A, 0W	-0.0090 kW ~ 0.0090 kW (0.0000 kW) (0.0000 kW)	
			1000V /500A	+1000V, +500A, +500kW	496.00 kW ~ 504.00 kW (500.02 kW) (500.17 kW)
				-1000V, -500A, +500kW	496.00 kW ~ 504.00 kW (499.96 kW) (499.93 kW)

SAMPLE

検査成績表
<TEST REPORT>

製造番号<Serial No.>
検査年月日<Test Date>

(No. 130501925)
(2021-05-25)
<YYYY-MM-DD>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input> <Volt/Curr/Pow/Freq/PF>	許容範囲 <Tolerance>	校正値<Calibration Value> *1	
					CH1	CH2
-8. 有効電力 力率 1 外部センサ TYPE 2 CT比: 1<Power PF=1 External Sensor TYPE 2 CT Ratio: 1>						
	AC+DC	150V /20A	150V, 20A, 3kW, 55Hz, PF=1	2.9910 kW ~ 3.0090 kW (2.9999 kW) (2.9999 kW)
			150V, 10A, 1.5kW, 55Hz, PF=1	1.4955 kW ~ 1.5045 kW (1.4999 kW) (1.5000 kW)
			150V, 2A, 0.3kW, 55Hz, PF=1	0.2964 kW ~ 0.3036 kW (0.3001 kW) (0.3002 kW)
			150V, 0.2A, 0.03kW, 55Hz, PF=1	0.0270 kW ~ 0.0330 kW (0.0300 kW) (0.0300 kW)
			150V(55Hz), 0A, 0W, -,-	-0.0030 kW ~ 0.0030 kW (0.0000 kW) (0.0000 kW)
		1000V /50A	1000V, 50A, 50kW, 55Hz, PF=1	49.850 kW ~ 50.150 kW (50.020 kW) (50.021 kW)
		1000V /10A	1000V, 10A, 10kW, 55Hz, PF=1	9.970 kW ~ 10.030 kW (10.003 kW) (10.003 kW)
		15V /50A	15V, 50A, 750W, 55Hz, PF=1	747.75 W ~ 752.25 W (750.33 W) (750.33 W)
		15V /10A	15V, 10A, 150W, 55Hz, PF=1	149.55 W ~ 150.45 W (150.03 W) (150.03 W)
			15V(55Hz), 0A, 0W, -,-	-0.15 W ~ 0.15 W (0.00 W) (0.00 W)
	DC	15V /10A	+15V, +10A, +150W	148.80 W ~ 151.20 W (150.06 W) (150.08 W)
			-15V, -10A, +150W	148.80 W ~ 151.20 W (149.97 W) (149.94 W)
			+15V, +0.1A, +1.5W	0.60 W ~ 2.40 W (1.55 W) (1.55 W)
			0V, 0A, 0W	-0.90 W ~ 0.90 W (0.00 W) (0.00 W)
		1000V /50A	+1000V, +50A, +50kW	49.600 kW ~ 50.400 kW (49.984 kW) (49.996 kW)
			-1000V, -50A, +50kW	49.600 kW ~ 50.400 kW (50.004 kW) (49.994 kW)

SAMPLE

検査成績表
<TEST REPORT>

製造番号<Serial No.> (No. 130501925)
 検査年月日<Test Date> (2021-05-25)
 <YYYY-MM-DD>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1		
			<Volt/Curr/Pow/Freq/PF>			CH1	CH2	
-9. 有効電力 力率 0 外部センサ TYPE 1 CT比:10<Power PF=0 External Sensor TYPE 1 CT Ratio:10>								
AC+DC		150V /200A	150V, 200A, 0W,	55Hz, LEAD 0	-0.045 kW ~ 0.045 kW (-0.001 kW) (-0.001 kW)	
			150V, 200A, 0W,	55Hz, LAG 0	-0.045 kW ~ 0.045 kW (0.001 kW) (0.001 kW)	
		15V /500A	15V, 500A, 0W,	55Hz, LEAD 0	-0.0112 kW ~ 0.0112 kW (0.0000 kW) (0.0000 kW)	
			15V, 500A, 0W,	55Hz, LAG 0	-0.0112 kW ~ 0.0112 kW (0.0000 kW) (0.0000 kW)	

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1		
			<Volt/Curr/Pow/Freq/PF>			CH1	CH2	
-10. 有効電力 力率 0 外部センサ TYPE 2 CT比:1<Power PF=0 External Sensor TYPE 2 CT Ratio:1>								
AC+DC		150V /20A	150V, 20A, 0W,	55Hz, LEAD 0	-0.0045 kW ~ 0.0045 kW (0.0000 kW) (0.0000 kW)	
			150V, 20A, 0W,	55Hz, LAG 0	-0.0045 kW ~ 0.0045 kW (0.0000 kW) (0.0001 kW)	
		15V /50A	15V, 50A, 0W,	55Hz, LEAD 0	-1.12 W ~ 1.12 W (0.00 W) (0.00 W)	
			15V, 50A, 0W,	55Hz, LAG 0	-1.12 W ~ 1.12 W (-0.01 W) (0.00 W)	

2. ピーク値測定<Peak Measurement>

項目 <Item>	レクティファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 **A <Tolerance>	校正値<Calibration Value> *1, **A	
			<Volt/Curr>	<Freq>		CH1	CH2
-1. 電圧<Voltage>							
AC+DC		3.6kV	600V	55Hz	±(0.7766 kV ~ 0.9205 kV) (0.8491 kV) (0.8491 kV)
-2. 電流<Current>							
AC+DC		60A	10A	55Hz	±(12.943 A ~ 15.342 A) (14.140 A) (14.144 A)
-3. 外部センサ TYPE 1 CT比:10<External Sensor TYPE 1 CT Ratio:10>							
AC+DC		3kA	500A	55Hz	±(0.6450 kA ~ 0.7692 kA) (0.7069 kA) (0.7070 kA)
-4. 外部センサ TYPE 2 CT比:1<External Sensor TYPE 2 CT Ratio:1>							
AC+DC		60A	10A	55Hz	±(12.900 A ~ 15.384 A) (14.176 A) (14.181 A)

**A. 判定は校正値と同じ符号の許容範囲を用いて行っています。

<The decision depends on the tolerance range of the sign same as the calibration value.>

SAMPLE

検査成績表
(TEST REPORT)

製造番号<Serial No.>
検査年月日<Test Date>

(No. 130501925)
(2021-05-25)
<YYYY-MM-DD>

3. 周波数特性<Frequency Characteristic>

項目 <Item>	レクティ ファイア <Rect>	ゼロ クロス フィルタ <Filter>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1	
				<Volt/Curr/ Pow/PF>	<Freq>		CH1	CH2
-1. 電圧<Voltage>								
	AC+DC	500Hz	15V	15V	10Hz	14.955 V ~ 15.045 V	(14.998 V)	(14.998 V)
					16Hz	14.970 V ~ 15.030 V	(15.000 V)	(14.999 V)
					50Hz	14.978 V ~ 15.022 V	(15.000 V)	(15.000 V)
					60Hz	14.978 V ~ 15.022 V	(15.000 V)	(15.000 V)
					500Hz	14.970 V ~ 15.030 V	(14.999 V)	(14.999 V)
		200kHz			10kHz	14.955 V ~ 15.045 V	(14.999 V)	(14.999 V)
					50kHz	14.880 V ~ 15.120 V	(14.999 V)	(14.999 V)
					100kHz	14.640 V ~ 15.360 V	(14.999 V)	(14.999 V)
-2. 電流<Current>								
	AC+DC	500Hz	5A	5A	10Hz	4.9850 A ~ 5.0150 A	(5.0010 A)	(5.0011 A)
					16Hz	4.9900 A ~ 5.0100 A	(5.0004 A)	(5.0005 A)
					50Hz	4.9925 A ~ 5.0075 A	(5.0007 A)	(5.0007 A)
					60Hz	4.9925 A ~ 5.0075 A	(5.0005 A)	(5.0005 A)
					500Hz	4.9900 A ~ 5.0100 A	(5.0004 A)	(5.0005 A)
		200kHz			1kHz	4.9850 A ~ 5.0150 A	(5.0009 A)	(5.0010 A)
					5kHz	4.9710 A ~ 5.0290 A	(5.0089 A)	(5.0087 A)
			2A	2A	100kHz	1.9080 A ~ 2.0920 A	(2.0055 A)	(2.0094 A)
-3. 有効電力<Power>								
	AC+DC	500Hz	15V /5A	15V, 5A, 75W, PF=1	10Hz	74.775 W ~ 75.225 W	(75.002 W)	(75.001 W)
					16Hz	74.850 W ~ 75.150 W	(75.003 W)	(75.004 W)
					50Hz	74.888 W ~ 75.112 W	(75.010 W)	(75.009 W)
					60Hz	74.888 W ~ 75.112 W	(75.009 W)	(75.010 W)
					500Hz	74.850 W ~ 75.150 W	(75.011 W)	(75.011 W)
		200kHz			1kHz	74.775 W ~ 75.225 W	(75.003 W)	(75.004 W)
					5kHz	74.565 W ~ 75.435 W	(75.082 W)	(75.079 W)
-4. 外部センサ TYPE 1 CT比:10<External Sensor TYPE 1 CT Ratio:10>								
	AC+DC	500Hz	200A	200A	10Hz	199.20 A ~ 200.80 A	(200.03 A)	(200.02 A)
					16Hz	199.20 A ~ 200.80 A	(199.99 A)	(199.99 A)
					50Hz	199.40 A ~ 200.60 A	(199.99 A)	(199.99 A)
					60Hz	199.40 A ~ 200.60 A	(199.96 A)	(199.96 A)
					500Hz	199.20 A ~ 200.80 A	(199.98 A)	(199.99 A)
		200kHz			1kHz	199.00 A ~ 201.00 A	(199.98 A)	(199.99 A)
					10kHz	190.00 A ~ 210.00 A	(199.21 A)	(199.23 A)
-5. 外部センサ TYPE 2 CT比:1<External Sensor TYPE 2 CT Ratio:1>								
	AC+DC	500Hz	10A	10A	10Hz	9.960 A ~ 10.040 A	(10.001 A)	(10.001 A)
					16Hz	9.960 A ~ 10.040 A	(10.001 A)	(10.001 A)
					50Hz	9.970 A ~ 10.030 A	(10.001 A)	(10.001 A)
					60Hz	9.970 A ~ 10.030 A	(10.001 A)	(10.001 A)
					500Hz	9.960 A ~ 10.040 A	(9.999 A)	(9.999 A)
		200kHz			1kHz	9.950 A ~ 10.050 A	(9.999 A)	(10.000 A)
					10kHz	9.500 A ~ 10.500 A	(9.981 A)	(9.982 A)

SAMPLE

検査成績表
<TEST REPORT>

製造番号<Serial No.> (No. 130501925)
 検査年月日<Test Date> (2021-05-25)
 <YYYY-MM-DD>

項目 <Item>	レクティ ファイア <Rect>	ゼロ クロス フィルタ <Filter>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1	
				<Volt/Curr/ Pow/PF>	<Freq>		CH1	CH2
-6. 有効電力 外部センサ TYPE 1 CT比:10<Power External Sensor TYPE 1 CT Ratio:10>								
	AC+DC	500Hz	15V /200A	15V, 200A, 3kW, PF=1	10Hz	2.9880 kW ~ 3.0120 kW (2.9999 kW) (2.9998 kW)
					16Hz	2.9880 kW ~ 3.0120 kW (2.9999 kW) (2.9998 kW)
					50Hz	2.9910 kW ~ 3.0090 kW (3.0001 kW) (3.0000 kW)
					60Hz	2.9910 kW ~ 3.0090 kW (3.0006 kW) (3.0006 kW)
					500Hz	2.9880 kW ~ 3.0120 kW (2.9997 kW) (2.9997 kW)
		200kHz			1kHz	2.9850 kW ~ 3.0150 kW (2.9993 kW) (2.9994 kW)
					10kHz	2.8500 kW ~ 3.1500 kW (2.9841 kW) (2.9846 kW)
-7. 有効電力 外部センサ TYPE 2 CT比:1<Power External Sensor TYPE 2 CT Ratio:1>								
	AC+DC	500Hz	15V /10A	15V, 10A, 150W, PF=1	10Hz	149.40 W ~ 150.60 W (149.99 W) (149.99 W)
					16Hz	149.40 W ~ 150.60 W (149.99 W) (149.99 W)
					50Hz	149.55 W ~ 150.45 W (150.01 W) (150.01 W)
					60Hz	149.55 W ~ 150.45 W (150.04 W) (150.04 W)
					500Hz	149.40 W ~ 150.60 W (150.00 W) (150.00 W)
		200kHz			1kHz	149.25 W ~ 150.75 W (149.98 W) (149.98 W)
					10kHz	142.50 W ~ 157.50 W (149.38 W) (149.41 W)

4. 周波数測定<Frequency Measurement>

項目 <Item>	レクティ ファイア <Rect>	レンジ <Range>	入力<Input>		許容範囲 <Tolerance>	校正値<Calibration Value> *1	
			<Volt/Curr>	<Freq>		CH1	CH2
-1. 電圧, 電流<Voltage, Current>							
	AC+DC	15V /200mA	15V, 0A	55Hz	54.944 Hz ~ 55.056 Hz (55.000 Hz) (55.000 Hz)
			0V, 200mA	55Hz	54.944 Hz ~ 55.056 Hz (55.000 Hz) (55.000 Hz)
-2. 外部センサ TYPE 1 CT比:10<External Sensor TYPE 1 CT Ratio:10>							
	AC+DC	100A	100A	55Hz	54.944 Hz ~ 55.056 Hz (55.000 Hz) (55.000 Hz)
-3. 外部センサ TYPE 2 CT比:1<External Sensor TYPE 2 CT Ratio:1>							
	AC+DC	10A	10A	55Hz	54.944 Hz ~ 55.056 Hz (55.000 Hz) (55.000 Hz)