

Important Notice Concerning Past Recalls

Although the following products have been discontinued, please note this important announcement regarding their recall that was previously issued.

This notice serves to inform customers that this recall program has ended based on actual recall performance, length of recall period, etc.

Date of occurrence	Product Name and Number	Reason for Recall	Notes
	Shipped Period		
2003/6/4	Meter Relay 2101,2102	There was risk of electric shock due to the peeling of the sealant used for insulating the range knob.	Manufacture of products ended 2005/12/20
	1981/1/1 to 2003/5/5		

Contact: Quality Assurance Department
HIOKI E.E. Corporation
TEL: 0268-28-0582 FAX: 0268-28-0595
E-mail: hinsyou@hioki.co.jp