

SAMPLE
 検査成績表
 <TEST REPORT>

1/1

品名<Model Name> (バッテリーテスタ<BATTERY TESTER >)
 形名<Model Number> (BT3554)^{*1}
 製造番号<Serial No.> (No. 210180000)
 検査年月日<Test Date> (2021-01-14)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (24.6 °C, 39 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 ^{*2} <Calibration Value>
確度<Accuracy>				
抵抗<Resistance>	3mΩ	0mΩ	0.000 mΩ ~ 0.008 mΩ	(0.002 mΩ)
		1mΩ	0.982 mΩ ~ 1.018 mΩ	(1.000 mΩ)
		3mΩ	2.962 mΩ ~ 3.038 mΩ	(3.000 mΩ)
	30mΩ	0mΩ	0.00 mΩ ~ 0.06 mΩ	(0.00 mΩ)
		10mΩ	9.86 mΩ ~ 10.14 mΩ	(10.00 mΩ)
		30mΩ	29.70 mΩ ~ 30.30 mΩ	(30.00 mΩ)
	300mΩ	0mΩ	0.0 mΩ ~ 0.6 mΩ	(0.0 mΩ)
		100mΩ	98.6 mΩ ~ 101.4 mΩ	(100.0 mΩ)
		300mΩ	297.0 mΩ ~ 303.0 mΩ	(300.0 mΩ)
	3Ω	0Ω	0.000 Ω ~ 0.006 Ω	(0.000 Ω)
		1Ω	0.986 Ω ~ 1.014 Ω	(1.000 Ω)
		3Ω	2.970 Ω ~ 3.030 Ω	(3.000 Ω)
電圧<Volt>	6V	0V	-0.006 V ~ 0.006 V	(0.000 V)
		3V	2.992 V ~ 3.008 V	(3.000 V)
		5.95V	5.940 V ~ 5.960 V	(5.950 V)
	60V	30V	29.92 V ~ 30.08 V	(30.00 V)
		-30V	-30.08 V ~ -29.92 V	(-30.00 V)
		59.5V	59.40 V ~ 59.60 V	(59.50 V)
温度<Temperature>		0°C(459.29Ω)	-0.5 °C ~ 0.5 °C	(0.0 °C)
		25°C(500.00Ω)	24.5 °C ~ 25.5 °C	(25.0 °C)
		50°C(540.37Ω)	49.5 °C ~ 50.5 °C	(50.0 °C)

備考<Note>

- ・ * 1 製品本体表記に基づき、形名欄への記載形名を以下のようにしています。<The model number stated in this document correspond to the respective model number indicated on the instrument as listed below:> BT3554-10: BT3554 BT3554-11: BT3554-01
- ・ * 2 FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()

No.BT3554-4