

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (バッテリーテスタ<BATTERY TESTER >)  
 形名<Model Number> ( BT3554-50 )<sup>\*1</sup>  
 製造番号<Serial No.> ( No. 210128781 )  
 検査年月日<Test Date> ( 2021-01-20 )  
 <YYYY-MM-DD>  
 検査条件<Test Conditions> ( 21.5 °C, 30 %RH )

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <sup>*2</sup> <Calibration Value>
確度<Accuracy>				
抵抗<Resistance>				
	3mΩ	0mΩ	0.000 mΩ ~ 0.008 mΩ	( 0.002 mΩ )
		1mΩ	0.982 mΩ ~ 1.018 mΩ	( 1.001 mΩ )
		3mΩ	2.962 mΩ ~ 3.038 mΩ	( 3.000 mΩ )
	30mΩ	0mΩ	0.00 mΩ ~ 0.06 mΩ	( 0.00 mΩ )
		10mΩ	9.86 mΩ ~ 10.14 mΩ	( 10.00 mΩ )
		30mΩ	29.70 mΩ ~ 30.30 mΩ	( 30.00 mΩ )
	300mΩ	0mΩ	0.0 mΩ ~ 0.6 mΩ	( 0.0 mΩ )
		100mΩ	98.6 mΩ ~ 101.4 mΩ	( 100.0 mΩ )
		300mΩ	297.0 mΩ ~ 303.0 mΩ	( 300.0 mΩ )
	3Ω	0Ω	0.000 Ω ~ 0.006 Ω	( 0.000 Ω )
		1Ω	0.986 Ω ~ 1.014 Ω	( 1.000 Ω )
		3Ω	2.970 Ω ~ 3.030 Ω	( 3.000 Ω )
電圧<Volt>				
	6V	0V	-0.006 V ~ 0.006 V	( 0.000 V )
		3V	2.992 V ~ 3.008 V	( 3.000 V )
		5.95V	5.940 V ~ 5.960 V	( 5.950 V )
	60V	30V	29.92 V ~ 30.08 V	( 30.00 V )
		-30V	-30.08 V ~ -29.92 V	( -30.00 V )
		59.5V	59.40 V ~ 59.60 V	( 59.50 V )
温度<Temperature>				
		0°C(459.29Ω)	-0.5 °C ~ 0.5 °C	( 0.0 °C )
		25°C(500.00Ω)	24.5 °C ~ 25.5 °C	( 25.0 °C )
		50°C(540.37Ω)	49.5 °C ~ 50.5 °C	( 50.0 °C )

備考<Note>

・ \*1 製品本体表記に基づき、形名欄への記載形名を以下のようにしています。<The model number stated in this document correspond to the respective model number indicated on the instrument as listed below:>

BT3554-50,BT3554-51,BT3554-52,BT3554-91,BT3554-92 : BT3554-50

・ \*2 FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
( PASS )	( )	( )